



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB85917

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4	02/20/2017	14:00
2	I1884-14		1	>100.0	>212.0	02/20/2017	14:32
3	I1884-14DUP		1	>100.0	>212.0	02/20/2017	15:05
4	I1884-17		1	>100.0	>212.0	02/20/2017	15:30
5	I1884-20		1	>100.0	>212.0	02/20/2017	15:55

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 85917

WorkList Name : FP22017

WorkList ID : 95786

Date : 2/20/2017 1:25:25 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/21/2017	Water	11884-14 B	Flash Point		PSEG01		MANHOLE-001	02/16/2017	1010A
02/21/2017	Water	11884-17 C	Flash Point		PSEG01		MANHOLE-002	02/16/2017	1010A
02/21/2017	Water	11884-20 B	Flash Point		PSEG01		GROUNDWATER	02/16/2017	1010A

Date/Time 2/20/17 1:30 PM
 Received by: KS
 Relinquished by: [Signature]

Date/Time 02/20/17 16:00
 Received by: [Signature]
 Relinquished by: KS