


7470A

UB 85928
INSTRUMENT ID: CVL

Linear

 μ Abs.:

36887



A= 0.0000e+000

B= 2.7000e-004

C= 3.1732e-002

Rho= 0.9999048

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.025	0.025	-24	0.000	-24				
0.2	0.200	0.223	0.023	710	0.0 %	710				
2.5	2.500	2.397	-0.103	8759	0.0 %	8759				
5.0	5.000	5.060	0.060	18623	0.0 %	18623				
7.5	7.500	7.504	0.004	27674	0.0 %	27674				
10.0	10.000	9.991	-0.009	36887	0.0 %	36887				

LB85928 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Feb 2017 10:36:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-24	-	0.0000	7470A	PPB	21 Feb 2017 10:56:51	S	Std
0.2 - 1	50.2	710	-	0.2000	7470A	PPB	21 Feb 2017 10:58:59	S	Std
2.5 - 1	52.5	8759	-	2.5000	7470A	PPB	21 Feb 2017 11:01:08	S	Std
5.0 - 1	55	18623	-	5.0000	7470A	PPB	21 Feb 2017 11:03:16	S	Std
7.5 - 1	57.5	27674	-	7.5000	7470A	PPB	21 Feb 2017 11:05:24	S	Std
10.0 - 1	510	36887	-	10.0000	7470A	PPB	21 Feb 2017 11:10:05	S	Std
ICV40 - 1	ICV40	14545	3.9588	-	7470A	PPB	21 Feb 2017 11:12:55	U	SMPL
ICB40 - 1	ICB40	-47	0.0190	-	7470A	PPB	21 Feb 2017 11:15:01	U	SMPL
CCV50 - 1	CCV50	18109	4.9211	-	7470A	PPB	21 Feb 2017 11:17:07	U	SMPL
CCB50 - 1	CCB50	-56	0.0166	-	7470A	PPB	21 Feb 2017 11:19:14	U	SMPL
CRA - 1	CRA	715	0.2248	-	7470A	PPB	21 Feb 2017 11:21:20	U	SMPL
HighStd - 1	HighStd	36578	9.9077	-	7470A	PPB	21 Feb 2017 11:23:27	U	SMPL
ChkStd - 1	ChkStd	25711	6.9736	-	7470A	PPB	21 Feb 2017 11:25:34	U	SMPL
PB97076BL - 1	PBW	-64	0.0145	-	7470A	PPB	21 Feb 2017 11:27:41	U	SMPL
PB97076BS - 1	LCSW	14365	3.9102	-	7470A	PPB	21 Feb 2017 11:29:48	U	SMPL
PB97076TB - 1	PB96996TB	-52	0.0177	-	7470A	PPB	21 Feb 2017 11:31:56	U	SMPL
I1865-02 - 1	TILCON-SCREENINGS	-26	0.0247	-	7470A	PPB	21 Feb 2017 11:34:04	U	SMPL
I1869-02 - 1	TILCON-NO-57	-20	0.0263	-	7470A	PPB	21 Feb 2017 11:36:12	U	SMPL
I1878-07 - 1	B1-2	-10	0.0290	-	7470A	PPB	21 Feb 2017 11:38:21	U	SMPL
I1878-08 - 1	B1-5	3	0.0325	-	7470A	PPB	21 Feb 2017 11:40:28	U	SMPL
CCV51 - 1	CCV51	18245	4.9578	-	7470A	PPB	21 Feb 2017 11:42:34	U	SMPL
CCB51 - 1	CCB51	-50	0.0182	-	7470A	PPB	21 Feb 2017 11:44:41	U	SMPL
I1884-03 - 1	PIT-A-COMP	9	0.0342	-	7470A	PPB	21 Feb 2017 11:46:48	U	SMPL
I1884-06 - 1	PIT-B-COMP	-19	0.0266	-	7470A	PPB	21 Feb 2017 11:48:54	U	SMPL
I1884-09 - 1	PIT-C-COMP	-19	0.0266	-	7470A	PPB	21 Feb 2017 11:51:01	U	SMPL
I1884-12 - 1	PIT-D-COMP	-16	0.0274	-	7470A	PPB	21 Feb 2017 11:53:08	U	SMPL
I1884-14 - 1	MANHOLE-001	-23	0.0255	-	7470A	PPB	21 Feb 2017 11:55:16	U	SMPL
I1884-17 - 1	MANHOLE-002	-25	0.0250	-	7470A	PPB	21 Feb 2017 11:57:24	U	SMPL
I1884-20 - 1	GROUNDWATER	-1	0.0315	-	7470A	PPB	21 Feb 2017 11:59:31	U	SMPL
I1892-06 - 1	TP22-COMP16	17	0.0363	-	7470A	PPB	21 Feb 2017 12:01:40	U	SMPL
I1913-05 - 1	SP-ABCD-COMP	-7	0.0298	-	7470A	PPB	21 Feb 2017 12:03:48	U	SMPL
I1913-10 - 1	SP-EFGH-COMP	-23	0.0255	-	7470A	PPB	21 Feb 2017 12:05:54	U	SMPL
CCV52 - 1	CCV52	18502	5.0272	-	7470A	PPB	21 Feb 2017 12:08:01	U	SMPL
CCB52 - 1	CCB52	-50	0.0182	-	7470A	PPB	21 Feb 2017 12:10:08	U	SMPL
I1913-10DUP - 1	SP-EFGH-COMPDUP	-18	0.0269	-	7470A	PPB	21 Feb 2017 12:12:15	U	SMPL
I1913-10MS - 1	SP-EFGH-COMPMS	13829	3.7655	-	7470A	PPB	21 Feb 2017 12:14:22	U	SMPL
I1913-10MSD - 1	SP-EFGH-COMPMSD	13646	3.7161	-	7470A	PPB	21 Feb 2017 12:16:29	U	SMPL
I1913-10LX5 - 1		-39	0.0212	-	7470A	PPB	21 Feb 2017 12:18:36	U	SMPL
I1913-10A - 1		14573	3.9664	-	7470A	PPB	21 Feb 2017 12:20:43	U	SMPL
CCV53 - 1	CCV53	18237	4.9557	-	7470A	PPB	21 Feb 2017 12:22:51	U	SMPL
CCB53 - 1	CCB53	-36	0.0220	-	7470A	PPB	21 Feb 2017 12:24:59	U	SMPL