



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85936

Temp.: 23.4 °C

BalanceID: WC SC-4

SLOPE.: 99.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/17/2017	15:00
2	CAL2	1	Water			7.00	02/17/2017	15:01
3	CAL3	1	Water			10.02	02/17/2017	15:02
4	ICV	1	Water			7.01	02/17/2017	15:03
5	CCV1	1	Water			2.01	02/17/2017	15:04
6	I1913-05	1	Solid	20.02	20	6.92	02/17/2017	15:05
7	I1913-10	1	Solid	20.05	20	7.86	02/17/2017	15:06
8	I1913-10DUP	1	Solid	20.03	20	7.87	02/17/2017	15:07
9	CCV2	1	Water			12.02	02/17/2017	15:08

LB85936

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I913 WorkList ID : 95816 Date : 2/17/2017 1:55:20 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/22/2017	Solid	I1913-05	Corrosivity	Cool 4 deg C	PSEG01	F42	SP-ABCD-COMP	02/17/2017	9045C
02/22/2017	Solid	I1913-10	Corrosivity	Cool 4 deg C	PSEG01	F42	SP-EFGH-COMP	02/17/2017	9045C

Date/Time 02-17-17 11:55 AM
 Received by: JP
 Relinquished by: JP

Date/Time 02-17-17 5:30 PM
 Received by: JP
 Relinquished by: JP