



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: JIGNESHDate: 02/21/2017Run Number: LB85940BalanceID: WC SC-3DATE & TIME IN: 02/21/2017 13:30DATE & TIME OUT: 02/21/2017 17:40OVEN TEMP: 105 °COVEN TEMP: 104 °C

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Weight (g)	Result mg/L
1	LB85940BL	LB85940BL	1.4632	1.4632	100	1.4632	1.4632	1.4632	0	0
2	LB85940BS	LB85940BS	1.3785	1.3785	100	1.4331	1.4331	1.4331	0.0546	546
3	I1889-01	002M-35THAVE (FEB)	1.418	1.418	400	1.4291	1.4291	1.4291	0.0111	27.8
4	I1889-01DUP	002M-35THAVE (FEB) DUP	1.4244	1.4244	400	1.4353	1.4353	1.4353	0.0109	27.3

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

1385940

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS I1889 WorkList ID : 95817 Date : 2/21/2017 1:05:21 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1889-01	TSS	Cool 4 deg C	TULL01	I33	002M-35THAVE(FEB)	02/20/2017	SM2540 D

Date/Time 02-21-17 11:05 PM
Received by: JP
Relinquished by: _____

Date/Time 02-21-17 2:00 PM
Received by: CP
Relinquished by: JD