



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB85942

Temp.: 24.2 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/21/2017	13:35
2	CAL2	1	Water			7.02	02/21/2017	13:36
3	CAL3	1	Water			10.03	02/21/2017	13:39
4	ICV	1	Water			6.99	02/21/2017	13:40
5	CCV1	1	Water			2.02	02/21/2017	13:41
6	I1929-01	1	Solid	20.03	20	8.24	02/21/2017	13:42
7	I1929-01DUP	1	Solid	20.05	20	8.25	02/21/2017	13:43
8	CCV2	1	Water			12.02	02/21/2017	13:45

2385942

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I1929

WorkList ID : 95818

Date : 2/21/2017 1:21:38 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I1929-01	pH	Cool 4 deg C	TULL01	I31	GRIT-STOCKPILE	02/20/2017	9045C

Date/Time 02-21-17 1:21 PM
Received by: JP
Relinquished by: JP

Date/Time 02-21-17 5:00 PM
Received by: JP
Relinquished by: JP