


7471A

LB 85947
INSTRUMENT ID: CV1 μ Abs.:

47607



Linear

A= 0.0000e+000

B= 2.1097e-004

C= -9.8511e-003

Rho= 0.9998778

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.009	-0.009	4	0.000	4				
0.2	0.200	0.188	-0.012	940	0.0 %	940				
2.5	2.500	2.509	0.009	11940	0.0 %	11940				
5.0	5.000	5.084	0.084	24146	0.0 %	24146				
7.5	7.500	7.393	-0.107	35090	0.0 %	35090				
10.0	10.000	10.034	0.034	47607	0.0 %	47607				

LB85947 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 21 Feb 2017 14:36:41

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	4	-	0.0000	7471A	PPB	21 Feb 2017 14:39:03	S	Std
0.2 - 1	50.2	940	-	0.2000	7471A	PPB	21 Feb 2017 14:41:21	S	Std
2.5 - 1	52.5	11940	-	2.5000	7471A	PPB	21 Feb 2017 14:43:38	S	Std
5.0 - 1	55	24146	-	5.0000	7471A	PPB	21 Feb 2017 14:45:56	S	Std
7.5 - 1	57.5	35090	-	7.5000	7471A	PPB	21 Feb 2017 14:48:13	S	Std
10.0 - 1	510	47607	-	10.0000	7471A	PPB	21 Feb 2017 14:50:29	S	Std
ICV41 - 1	ICV41	18709	3.9372	-	7471A	PPB	21 Feb 2017 14:53:15	U	SMPL
ICB41 - 1	ICB41	-35	-0.0172	-	7471A	PPB	21 Feb 2017 14:55:30	U	SMPL
CCV54 - 1	CCV54	23719	4.9942	-	7471A	PPB	21 Feb 2017 14:57:46	U	SMPL
CCB54 - 1	CCB54	-34	-0.0170	-	7471A	PPB	21 Feb 2017 15:00:02	U	SMPL
CRA - 1	CRA	920	0.1842	-	7471A	PPB	21 Feb 2017 15:02:17	U	SMPL
HighStd - 1	HighStd	46879	9.8803	-	7471A	PPB	21 Feb 2017 15:04:33	U	SMPL
ChkStd - 1	ChkStd	33538	7.0657	-	7471A	PPB	21 Feb 2017 15:06:49	U	SMPL
PB97078BL - 1	PBS	-33	-0.0168	-	7471A	PPB	21 Feb 2017 15:09:06	U	SMPL
PB97078BS - 1	LCSS	16441	3.4587	-	7471A	PPB	21 Feb 2017 15:11:23	U	SMPL
I1915-03 - 1	B1-1	74060	15.6148	-	7471A	PPB	21 Feb 2017 15:13:40	U	SMPL
I1923-01 - 1	CORBETT-BF	171	0.0262	-	7471A	PPB	21 Feb 2017 15:15:57	U	SMPL
I1926-01 - 1	CORBETT-SE	275	0.0482	-	7471A	PPB	21 Feb 2017 15:18:15	U	SMPL
I1926-01DUP - 1	CORBETT-SEDUP	298	0.0530	-	7471A	PPB	21 Feb 2017 15:20:32	U	SMPL
I1926-01MS - 1	CORBETT-SEMS	17058	3.5889	-	7471A	PPB	21 Feb 2017 15:22:48	U	SMPL
CCV55 - 1	CCV55	24030	5.0598	-	7471A	PPB	21 Feb 2017 15:25:03	U	SMPL
CCB55 - 1	CCB55	-31	-0.0164	-	7471A	PPB	21 Feb 2017 15:27:19	U	SMPL
I1926-01MSD - 1	CORBETT-SEMSD	16772	3.5286	-	7471A	PPB	21 Feb 2017 15:29:34	U	SMPL
I1926-01LX5 - 1		32	-0.0031	-	7471A	PPB	21 Feb 2017 15:31:50	U	SMPL
I1926-01A - 1		18560	3.9058	-	7471A	PPB	21 Feb 2017 15:34:06	U	SMPL
I1915-03DLX5 - 1		14924	3.1387	-	7471A	PPB	21 Feb 2017 15:36:23	U	SMPL
CCV56 - 1	CCV56	23893	5.0309	-	7471A	PPB	21 Feb 2017 15:38:39	U	SMPL
CCB56 - 1	CCB56	-22	-0.0145	-	7471A	PPB	21 Feb 2017 15:40:56	U	SMPL