



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85950

Temp.: 24.1 °C

BalanceID: WC SC-4

SLOPE.: 99.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/21/2017	17:30
2	CAL2	1	Water			7.02	02/21/2017	17:31
3	CAL3	1	Water			10.02	02/21/2017	17:33
4	ICV	1	Water			7.01	02/21/2017	17:35
5	CCV1	1	Water			2.02	02/21/2017	17:36
6	I1935-02	1	Solid	20.02	20	8.10	02/21/2017	17:37
7	I1935-02DUP	1	Solid	20.04	20	8.11	02/21/2017	17:39
8	I1936-03	1	Solid	20.03	20	8.31	02/21/2017	17:40
9	I1936-04	1	Solid	20.05	20	8.51	02/21/2017	17:41
10	CCV2	1	Water			12.01	02/21/2017	17:43

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I1935

WorkList ID : 95826

Date : 2/21/2017 5:22:58 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I1935-02	Corrosivity	Cool 4 deg C	ENTA04	I31	WC-B31-B34-001	02/21/2017	9045C
	Solid	I1936-03	Corrosivity	Cool 4 deg C	ENTA04	I32	WC-B38-B39-001	02/21/2017	9045C
	Solid	I1936-04	Corrosivity	Cool 4 deg C	ENTA04	I32	WC-B55-B63-001	02/21/2017	9045C

Date/Time 02-21-17 5:22 PM
 Received by: JD
 Relinquished by: JD

Date/Time 02-21-17 6:30 PM
 Received by: CP
 Relinquished by: JD