


245.1

LB 85955
INSTRUMENT ID: CV2

Linear

 μ Abs.:

33232



A= 0.0000e+000

B= 2.9944e-004

C= -1.3490e-002

Rho= 0.9999412

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.019	-0.019	-17	0.000	-17				
0.2	0.200	0.178	-0.022	638	0.0 %	638				
2.5	2.500	2.522	0.022	8469	0.0 %	8469				
5.0	5.000	5.018	0.018	16802	0.0 %	16802				
7.5	7.500	7.563	0.063	25304	0.0 %	25304				
10.0	10.000	9.937	-0.063	33232	0.0 %	33232				

LB85955 INSTRUMENT IO : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 22 Feb 2017 10:55:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-17	-	0.0000	245.1	PPB	22 Feb 2017 11:32:05	S	Std
0.2 - 1	50.2	638	-	0.2000	245.1	PPB	22 Feb 2017 11:34:13	S	Std
2.5 - 1	52.5	8469	-	2.5000	245.1	PPB	22 Feb 2017 11:36:22	S	Std
5.0 - 1	55	16802	-	5.0000	245.1	PPB	22 Feb 2017 11:38:30	S	Std
7.5 - 1	57.5	25304	-	7.5000	245.1	PPB	22 Feb 2017 11:40:38	S	Std
10.0 - 1	510	33232	-	10.0000	245.1	PPB	22 Feb 2017 11:42:45	S	Std
ICV42 - 1	ICV42	13203	3.9400	-245.1		PPB	22 Feb 2017 11:45:31	U	SMPL
ICB42 - 1	ICB42	-38	-0.0249	-245.1		PPB	22 Feb 2017 11:47:37	U	SMPL
CCV57 - 1	CCV57	16596	4.9560	-245.1		PPB	22 Feb 2017 11:49:44	U	SMPL
CCB57 - 1	CCB57	-41	-0.0258	-245.1		PPB	22 Feb 2017 11:51:50	U	SMPL
CRA - 1	CRA	635	0.1767	-245.1		PPB	22 Feb 2017 11:53:57	U	SMPL
HighStd - 1	HighStd	32650	9.7632	-245.1		PPB	22 Feb 2017 11:56:04	U	SMPL
ChkStd - 1	ChkStd	24016	7.1778	-245.1		PPB	22 Feb 2017 11:58:11	U	SMPL
PB97087BL - 1	PBW	-58	-0.0309	-245.1		PPB	22 Feb 2017 12:00:18	U	SMPL
PB97087BS - 1	LCSW	13612	4.0625	-245.1		PPB	22 Feb 2017 12:02:26	U	SMPL
1828-02 - 1	WINTER-2017	15745	4.7012	-245.1		PPB	22 Feb 2017 12:04:33	U	SMPL
1828-02DUP - 1	WINTER-2017DUP	15357	4.5850	-245.1		PPB	22 Feb 2017 12:06:42	U	SMPL
1828-02MS - 1	WINTER-2017MS	28415	8.4950	-245.1		PPB	22 Feb 2017 12:11:26	U	SMPL
1828-02MSD - 1	WINTER-2017MSD	28899	8.6400	-245.1		PPB	22 Feb 2017 12:13:35	U	SMPL
CCV58 - 1	CCV58	16513	4.9311	-245.1		PPB	22 Feb 2017 12:15:42	U	SMPL
CCB58 - 1	CCB58	-55	-0.0300	-245.1		PPB	22 Feb 2017 12:17:49	U	SMPL
1828-02LX5 - 1		3584	1.0597	-245.1		PPB	22 Feb 2017 12:19:56	U	SMPL
1828-02A - 1		28661	8.5687	-245.1		PPB	22 Feb 2017 12:24:38	U	SMPL
CCV59 - 1	CCV59	16426	4.9051	-245.1		PPB	22 Feb 2017 12:26:45	U	SMPL
CCB59 - 1	CCB59	-47	-0.0276	-245.1		PPB	22 Feb 2017 12:28:52	U	SMPL