



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB85960

Temp.: 21.3 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/22/2017	11:00
2	CAL2	1	Water			7.00	02/22/2017	11:01
3	CAL3	1	Water			10.03	02/22/2017	11:02
4	ICV	1	Water			7.02	02/22/2017	11:03
5	CCV1	1	Water			2.01	02/22/2017	11:05
6	I1938-01	1	Water			4.35	02/22/2017	11:06
7	I1938-02	1	Water			4.53	02/22/2017	11:07
8	I1938-03	1	Water			4.60	02/22/2017	11:10
9	I1945-01	1	Water			5.55	02/22/2017	11:11
10	I1945-01DUP	1	Water			5.56	02/22/2017	11:12
11	CCV2	1	Water			12.01	02/22/2017	11:13

LB85960

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I1938

WorkList ID : 95842

Date : 2/22/2017 10:50:05 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1938-01	pH	Cool 4 deg C	STER13	I32	CELL-A-POOL-WATER	02/21/2017	SM 4500-PH B
	Water	I1938-02	pH	Cool 4 deg C	STER13	I32	CELL-B-POOL-WATER	02/21/2017	SM 4500-PH B
	Water	I1938-03	pH	Cool 4 deg C	STER13	I32	CELL-C-POOL-WATER	02/21/2017	SM 4500-PH B
	Water	I1945-01	pH	Cool 4 deg C	STER01	I13	POOLWATER	02/21/2017	SM 4500-PH B

Date/Time 02/22/17 10.50 AM

Received by: JP

Relinquished by: CA

Date/Time 02/22/17 12:20 PM

Received by: OP

Relinquished by: JP