


7471A

LB 85966
INSTRUMENT ID: CV2

Linear

 μ Abs.:

48217



A= 0.0000e+000

B= 2.0782e-004

C= 2.0722e-002

Rho= 0.9998793

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.024	0.024	15	0.000	15				
0.2	0.200	0.207	0.007	896	0.0 %	896				
2.5	2.500	2.440	-0.060	11643	0.0 %	11643				
5.0	5.000	5.077	0.077	24328	0.0 %	24328				
7.5	7.500	7.411	-0.089	35559	0.0 %	35559				
10.0	10.000	10.041	0.041	48217	0.0 %	48217				

LB85966 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 22 Feb 2017 14:35:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	80	15	-	0.0000	7471A	PPB	22 Feb 2017 14:36:08	S	Std
0.2 - 1	50.2	896	-	0.2000	7471A	PPB	22 Feb 2017 14:38:26	S	Std
2.5 - 1	52.5	11643	-	2.5000	7471A	PPB	22 Feb 2017 14:40:43	S	Std
5.0 - 1	55	24328	-	5.0000	7471A	PPB	22 Feb 2017 14:43:01	S	Std
7.5 - 1	57.5	35559	-	7.5000	7471A	PPB	22 Feb 2017 14:45:18	S	Std
10.0 - 1	90	48217	-	10.0000	7471A	PPB	22 Feb 2017 14:47:34	S	Std
ICV43 - 1	ICV43	18743	3.9160	-7471A	PPB	PPB	22 Feb 2017 14:50:17	U	SMPL
ICB43 - 1	ICB43	-1	0.0205	-7471A	PPB	PPB	22 Feb 2017 14:52:32	U	SMPL
CCV60 - 1	CCV60	23890	4.9856	-7471A	PPB	PPB	22 Feb 2017 14:54:48	U	SMPL
CCB60 - 1	CCB60	-14	0.0178	-7471A	PPB	PPB	22 Feb 2017 14:57:04	U	SMPL
CRA - 1	CRA	892	0.2061	-7471A	PPB	PPB	22 Feb 2017 14:59:19	U	SMPL
HighStd - 1	HighStd	47980	9.9921	-7471A	PPB	PPB	22 Feb 2017 15:01:35	U	SMPL
ChkStd - 1	ChkStd	33714	7.0273	-7471A	PPB	PPB	22 Feb 2017 15:03:51	U	SMPL
PB97094BL - 1	PBS	-7	0.0193	-7471A	PPB	PPB	22 Feb 2017 15:06:07	U	SMPL
PB97094BS - 1	LCSS	16215	3.3906	-7471A	PPB	PPB	22 Feb 2017 15:08:24	U	SMPL
I1929-01 - 1	GRIT-STOCKPILE	4823	1.0231	-7471A	PPB	PPB	22 Feb 2017 15:10:41	U	SMPL
I1930-02 - 1	14A-9-11	343	0.0920	-7471A	PPB	PPB	22 Feb 2017 15:12:59	U	SMPL
I1931-01 - 1	P-2	160	0.0540	-7471A	PPB	PPB	22 Feb 2017 15:15:16	U	SMPL
I1931-03 - 1	P-3	164	0.0548	-7471A	PPB	PPB	22 Feb 2017 15:17:34	U	SMPL
I1931-05 - 1	P-4	202	0.0627	-7471A	PPB	PPB	22 Feb 2017 15:19:50	U	SMPL
CCV61 - 1	CCV61	24225	5.0553	-7471A	PPB	PPB	22 Feb 2017 15:22:06	U	SMPL
CCB61 - 1	CCB61	4	0.0216	-7471A	PPB	PPB	22 Feb 2017 15:24:22	U	SMPL
I1931-07 - 1	P-5	286	0.0802	-7471A	PPB	PPB	22 Feb 2017 15:26:38	U	SMPL
I1931-07DUP - 1	P-5DUP	238	0.0702	-7471A	PPB	PPB	22 Feb 2017 15:28:54	U	SMPL
I1931-07MS - 1	P-5MS	17418	3.6406	-7471A	PPB	PPB	22 Feb 2017 15:33:50	U	SMPL
I1931-07MSD - 1	P-5MSD	17282	3.6123	-7471A	PPB	PPB	22 Feb 2017 15:36:06	U	SMPL
I1931-09 - 1	P-6	172	0.0565	-7471A	PPB	PPB	22 Feb 2017 15:38:23	U	SMPL
I1931-11 - 1	P-7	267	0.0762	-7471A	PPB	PPB	22 Feb 2017 15:40:40	U	SMPL
I1931-07LX5 - 1		68	0.0349	-7471A	PPB	PPB	22 Feb 2017 15:42:57	U	SMPL
I1931-07A - 1		19081	3.9862	-7471A	PPB	PPB	22 Feb 2017 15:45:15	U	SMPL
CCV62 - 1	CCV62	24085	5.0262	-7471A	PPB	PPB	22 Feb 2017 15:47:33	U	SMPL
CCB62 - 1	CCB62	5	0.0218	-7471A	PPB	PPB	22 Feb 2017 15:49:49	U	SMPL