



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85968

Temp.: 24.3 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	02/22/2017	15:00
2	CAL2	1	Water			7.01	02/22/2017	15:01
3	CAL3	1	Water			10.02	02/22/2017	15:02
4	ICV	1	Water			7.00	02/22/2017	15:03
5	CCV1	1	Water			2.02	02/22/2017	15:05
6	I1948-06	1	Solid	20.02	20	10.06	02/22/2017	15:06
7	I1948-06DUP	1	Solid	20.04	20	10.08	02/22/2017	15:07
8	CCV2	1	Water			12.01	02/22/2017	15:10

LB 85968

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 11948 WorkList ID : 95850 Date : 2/22/2017 2:42:50 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I1948-06	Corrosivity	Cool 4 deg C	DEWB01	I31	TP23-24-COMP17	02/22/2017	9045C

Date/Time 02/22/17 2:42 PM
Received by: JH
Relinquished by: CL

Date/Time 02/22/17 3:50 PM
Received by: CS
Relinquished by: JD