



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB85970

Temp.: 24.3 °C

BalanceID: WC SC-4

SLOPE.: 98.9

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/22/2017	15:30
2	CAL2	1	Water			7.00	02/22/2017	15:31
3	CAL3	1	Water			10.02	02/22/2017	15:32
4	ICV	1	Water			7.01	02/22/2017	15:33
5	CCV1	1	Water			2.02	02/22/2017	15:34
6	I1952-01	1	Solid	20.04	20	9.02	02/22/2017	15:36
7	I1952-01DUP	1	Solid	20.02	20	9.03	02/22/2017	15:37
8	I1952-03	1	Solid	20.03	20	7.49	02/22/2017	15:38
9	CCV2	1	Water			12.02	02/22/2017	15:40

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I1952

WorkList ID : 95854

Date : 2/22/2017 3:20:46 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I1952-01	pH	Cool 4 deg C	GLOB02	I42	SPARTA-FINE-SAND	02/21/2017	9045C
	Solid	I1952-03	pH	Cool 4 deg C	GLOB02	I42	SPARTA-TOP-SOIL	02/21/2017	9045C

Date/Time 02/22/17 3:20 PM
 Received by: JP
 Relinquished by: CP

Date/Time 02/22/17 4:30 PM
 Received by: CP
 Relinquished by: JP