



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB85971

Temp.: 21.2 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	02/22/2017	15:40
2	CAL2	1	Water			7.01	02/22/2017	15:41
3	CAL3	1	Water			10.03	02/22/2017	15:42
4	ICV	1	Water			6.99	02/22/2017	15:44
5	CCV1	1	Water			2.01	02/22/2017	15:45
6	I1953-01	1	Water			9.04	02/22/2017	15:46
7	I1953-01DUP	1	Water			9.05	02/22/2017	15:47
8	CCV2	1	Water			12.02	02/22/2017	15:48

LB85971

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH11953 WorkList ID : 95855 Date : 2/22/2017 3:27:47 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1953-01	pH	Cool 4 deg C	DIST01	I31	ZAVE-01	02/22/2017	SM 4500-PH B

Date/Time 02/22/17 3:27 PM
 Received by: PH
 Relinquished by: CP

Date/Time 02/22/17 3:27 PM
 Received by: CP
 Relinquished by: JP