



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB85986

Temp.: 21.2 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	02/22/2017	15:40
2	CAL2	1	Water			7.01	02/22/2017	15:41
3	CAL3	1	Water			10.03	02/22/2017	15:42
4	ICV	1	Water			6.99	02/22/2017	15:43
5	CCV1	1	Water			2.01	02/22/2017	15:45
6	I1953-02	1	Water			9.04	02/22/2017	15:46
7	I1953-02DUP	1	Water			9.05	02/22/2017	15:47
8	I1956-01	1	Water			7.18	02/22/2017	15:49
9	CCV2	1	Water			12.01	02/22/2017	15:50

2B85486

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 11956

WorkList ID : 95861

Date : 2/22/2017 6:06:41 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/22/2017	Water	11953-02	pH		DIST01		2AVE-01	02/22/2017	SM 4500-PH B
02/22/2017	Water	11956-01	pH	Cool 4 deg C	PSEG01	F61	FRAC-TANK-A2622	02/22/2017	9045C

Date/Time 02/22/17 6:06 PM
Received by: JP
Relinquished by: [Signature]

Date/Time 02/22/17 6:40 PM
Received by: [Signature]
Relinquished by: SP