



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85987

Temp.: 21.2 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	02/22/2017	15:40
2	CAL2	1	Water			7.01	02/22/2017	15:41
3	CAL3	1	Water			10.03	02/22/2017	15:42
4	ICV	1	Water			6.99	02/22/2017	15:43
5	CCV1	1	Water			2.01	02/22/2017	15:43
6	I1956-02	1	Water			7.18	02/22/2017	15:49
7	I1956-02DUP	1	Water			7.19	02/22/2017	15:50
8	CCV2	1	Water			12.01	02/22/2017	15:51

2885987

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY i1956

WorkList ID : 95862

Date : 2/22/2017 6:08:12 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/22/2017	Water	I1956-02	Corrosivity	Cool 4 deg C	PSEG01	F61	FRAC-TANK-A2622	02/22/2017	9040C

Date/Time 02/22/17 6:08 PM
Received by: JS
Relinquished by: CS

Date/Time 02/22/17 6:40 PM
Received by: AP
Relinquished by: SP