


7470A

LB85989
INSTRUMENT ID: 041

Linear

 μ Abs.:

38010



A= 0.0000e+000

B= 2.6418e-004

C= 1.2820e-002

Rho= 0.9998085

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.011	0.011	-5	0.000	-5				
0.2	0.200	0.192	-0.008	679	0.0 %	679				
2.5	2.500	2.478	-0.022	9333	0.0 %	9333				
5.0	5.000	5.098	0.098	19250	0.0 %	19250				
7.5	7.500	7.366	-0.134	27833	0.0 %	27833				
10.0	10.000	10.054	0.054	38010	0.0 %	38010				

LB85989 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Feb 2017 11:36:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	-5	-	0.00007470A	PPB		23 Feb 2017 11:55:12	S	Std
0.2 - 1	SD.2	679	-	0.20007470A	PPB		23 Feb 2017 11:57:21	S	Std
2.5 - 1	SD.5	9333	-	2.50007470A	PPB		23 Feb 2017 11:59:29	S	Std
5.0 - 1	SD	19250	-	5.00007470A	PPB		23 Feb 2017 12:01:37	S	Std
7.5 - 1	SD.5	27833	-	7.50007470A	PPB		23 Feb 2017 12:03:45	S	Std
10.0 - 1	SD	38010	-	10.00007470A	PPB		23 Feb 2017 12:05:52	S	Std
ICV44 - 1	ICV44	14504	3.8444	-7470A	PPB		23 Feb 2017 12:08:27	U	SMPL
ICB44 - 1	ICB44	-102	-0.0141	-7470A	PPB		23 Feb 2017 12:10:33	U	SMPL
CCV63 - 1	CCV63	19103	5.0594	-7470A	PPB		23 Feb 2017 12:12:40	U	SMPL
CCB63 - 1	CCB63	-115	-0.0176	-7470A	PPB		23 Feb 2017 12:14:47	U	SMPL
CRA - 1	CRA	599	0.1711	-7470A	PPB		23 Feb 2017 12:16:53	U	SMPL
HighStd - 1	HighStd	37271	9.8589	-7470A	PPB		23 Feb 2017 12:19:00	U	SMPL
ChkStd - 1	ChkStd	25871	6.8473	-7470A	PPB		23 Feb 2017 12:21:07	U	SMPL
PB97117BL - 1	PBW	-119	-0.0186	-7470A	PPB		23 Feb 2017 12:23:14	U	SMPL
PB97117BS - 1	LCSW	13038	3.4571	-7470A	PPB		23 Feb 2017 12:25:22	U	SMPL
PB97117TB - 1	PB97090TB	-41	0.0020	-7470A	PPB		23 Feb 2017 12:27:30	U	SMPL
I1915-04 - 1	B1-1	2	0.0133	-7470A	PPB		23 Feb 2017 12:29:37	U	SMPL
I1929-02 - 1	GRIT-STOCKPILE	-8	0.0107	-7470A	PPB		23 Feb 2017 12:31:45	U	SMPL
I1935-02 - 1	WC-B31-B34-001	57	0.0279	-7470A	PPB		23 Feb 2017 12:33:54	U	SMPL
I1936-03 - 1	WC-B38-B39-001	-2	0.0123	-7470A	PPB		23 Feb 2017 12:36:01	U	SMPL
CCV64 - 1	CCV64	19034	5.0411	-7470A	PPB		23 Feb 2017 12:38:08	U	SMPL
CCB64 - 1	CCB64	-45	0.0009	-7470A	PPB		23 Feb 2017 12:40:15	U	SMPL
I1936-04 - 1	WC-B55-B63-001	4	0.0139	-7470A	PPB		23 Feb 2017 12:42:21	U	SMPL
I1936-04DUP - 1	WC-B55-B63-001DUP	-1	0.0126	-7470A	PPB		23 Feb 2017 12:44:28	U	SMPL
I1936-04MS - 1	WC-B55-B63-001MS	13820	3.6637	-7470A	PPB		23 Feb 2017 12:46:35	U	SMPL
I1936-04MSD - 1	WC-B55-B63-001MSD	13760	3.6479	-7470A	PPB		23 Feb 2017 12:48:42	U	SMPL
PB97118BL - 1	PBW	-9	0.0104	-7470A	PPB		23 Feb 2017 12:50:49	U	SMPL
PB97118BS - 1	LCSW	13608	3.6077	-7470A	PPB		23 Feb 2017 12:52:57	U	SMPL
PB97118TB - 1	PB97092TB	-32	0.0044	-7470A	PPB		23 Feb 2017 12:55:05	U	SMPL
I1930-04 - 1	14A-9-11	22	0.0186	-7470A	PPB		23 Feb 2017 12:57:13	U	SMPL
I1930-04DUP - 1	14A-9-11DUP	22	0.0186	-7470A	PPB		23 Feb 2017 12:59:22	U	SMPL
I1930-04MS - 1	14A-9-11MS	16281	4.3139	-7470A	PPB		23 Feb 2017 13:01:28	U	SMPL
CCV65 - 1	CCV65	18919	5.0108	-7470A	PPB		23 Feb 2017 13:03:35	U	SMPL
CCB65 - 1	CCB65	-43	0.0015	-7470A	PPB		23 Feb 2017 13:05:42	U	SMPL
I1930-04MSD - 1	14A-9-11MSD	15387	4.0777	-7470A	PPB		23 Feb 2017 13:07:49	U	SMPL
I1936-04LX5 - 1		-27	0.0057	-7470A	PPB		23 Feb 2017 13:09:56	U	SMPL
I1936-04A - 1		14348	3.8032	-7470A	PPB		23 Feb 2017 13:12:03	U	SMPL
I1930-04LX5 - 1		-22	0.0070	-7470A	PPB		23 Feb 2017 13:14:10	U	SMPL
I1930-04A - 1		14447	3.8294	-7470A	PPB		23 Feb 2017 13:16:18	U	SMPL
CCV66 - 1	CCV66	19036	5.0417	-7470A	PPB		23 Feb 2017 13:18:25	U	SMPL
CCB66 - 1	CCB66	-36	0.0033	-7470A	PPB		23 Feb 2017 13:20:33	U	SMPL