


7470A

LD 86006
INSTRUMENT ID: CY2

Linear

 μ Abs.:

37930



A= 0.0000e+000

B= 2.6459e-004

C= -6.7241e-003

Rho= 0.9999632

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.007	0.007	52	0.000	52				
0.2	0.200	0.203	0.003	793	0.0 %	793				
2.5	2.500	2.532	0.032	9596	0.0 %	9596				
5.0	5.000	4.937	-0.063	18684	0.0 %	18684				
7.5	7.500	7.492	-0.008	28340	0.0 %	28340				
10.0	10.000	10.029	0.029	37930	0.0 %	37930				

LB86006 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Feb 2017 08:49:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	52	-	0.0000	7470A	PPB	24 Feb 2017 08:55:52	S	Std
0.2 - 1	50.2	793	-	0.2000	7470A	PPB	24 Feb 2017 08:58:00	S	Std
2.5 - 1	52.5	9596	-	2.5000	7470A	PPB	24 Feb 2017 09:00:08	S	Std
5.0 - 1	55	18684	-	5.0000	7470A	PPB	24 Feb 2017 09:02:17	S	Std
7.5 - 1	57.5	28340	-	7.5000	7470A	PPB	24 Feb 2017 09:04:24	S	Std
10.0 - 1	510	37930	-	10.0000	7470A	PPB	24 Feb 2017 09:06:31	S	Std
ICV45 - 1	ICV45	13710	3.6208	-	7470A	PPB	24 Feb 2017 09:09:17	U	SMPL
ICB45 - 1	ICB45	5	-0.0054	-	7470A	PPB	24 Feb 2017 09:11:23	U	SMPL
CCV67 - 1	CCV67	18214	4.8125	-	7470A	PPB	24 Feb 2017 09:13:29	U	SMPL
CCB67 - 1	CCB67	8	-0.0046	-	7470A	PPB	24 Feb 2017 09:15:35	U	SMPL
CRA - 1	CRA	709	0.1809	-	7470A	PPB	24 Feb 2017 09:17:42	U	SMPL
HighStd - 1	HighStd	37017	9.7875	-	7470A	PPB	24 Feb 2017 09:19:48	U	SMPL
ChkStd - 1	ChkStd	25371	6.7061	-	7470A	PPB	24 Feb 2017 09:21:55	U	SMPL
PB97143BL - 1	PBW	-47	-0.0192	-	7470A	PPB	24 Feb 2017 09:24:03	U	SMPL
PB97143BS - 1	LCSW	14336	3.7864	-	7470A	PPB	24 Feb 2017 09:26:10	U	SMPL
PB97143TB - 1	PB97114TB	-37	-0.0165	-	7470A	PPB	24 Feb 2017 09:28:18	U	SMPL
I1948-06 - 1	TP23-24-COMP17	-7	-0.0086	-	7470A	PPB	24 Feb 2017 09:30:26	U	SMPL
I1955-04 - 1	B1-3	34	0.0023	-	7470A	PPB	24 Feb 2017 09:32:34	U	SMPL
I1955-08 - 1	B1-4	89	0.0168	-	7470A	PPB	24 Feb 2017 09:34:43	U	SMPL
I1956-02 - 1	FRAC-TANK-A2622	14	-0.0030	-	7470A	PPB	24 Feb 2017 09:36:50	U	SMPL
CCV68 - 1	CCV68	18350	4.8485	-	7470A	PPB	24 Feb 2017 09:38:57	U	SMPL
CCB68 - 1	CCB68	35	0.0025	-	7470A	PPB	24 Feb 2017 09:41:03	U	SMPL
I1956-02DUP - 1	FRAC-TANK-A2622DUP	6	-0.0051	-	7470A	PPB	24 Feb 2017 09:43:10	U	SMPL
I1956-02MS - 1	FRAC-TANK-A2622MS	13731	3.6263	-	7470A	PPB	24 Feb 2017 09:45:16	U	SMPL
I1956-02MSD - 1	FRAC-TANK-A2622MSD	13721	3.6237	-	7470A	PPB	24 Feb 2017 09:47:23	U	SMPL
PB97144BL - 1	PBW	23	-0.0006	-	7470A	PPB	24 Feb 2017 09:49:30	U	SMPL
PB97144BS - 1	LCSW	15085	3.9846	-	7470A	PPB	24 Feb 2017 09:51:37	U	SMPL
I1956-01 - 1	FRAC-TANK-A2622	-46	-0.0189	-	7470A	PPB	24 Feb 2017 09:53:45	U	SMPL
I1956-01DUP - 1	FRAC-TANK-A2622DUP	-22	-0.0125	-	7470A	PPB	24 Feb 2017 09:55:53	U	SMPL
I1956-01MS - 1	FRAC-TANK-A2622MS	14130	3.7319	-	7470A	PPB	24 Feb 2017 09:58:01	U	SMPL
I1956-01MSD - 1	FRAC-TANK-A2622MSD	13875	3.6644	-	7470A	PPB	24 Feb 2017 10:00:10	U	SMPL
I1957-13 - 1	FB20170222	-24	-0.0131	-	7470A	PPB	24 Feb 2017 10:02:16	U	SMPL
CCV69 - 1	CCV69	18670	4.9331	-	7470A	PPB	24 Feb 2017 10:04:23	U	SMPL
CCB69 - 1	CCB69	30	0.0012	-	7470A	PPB	24 Feb 2017 10:06:30	U	SMPL
I1956-02LX5 - 1		35	0.0025	-	7470A	PPB	24 Feb 2017 10:08:37	U	SMPL
I1956-02A - 1		15094	3.9870	-	7470A	PPB	24 Feb 2017 10:10:44	U	SMPL
I1956-01LX5 - 1		3	-0.0059	-	7470A	PPB	24 Feb 2017 10:12:51	U	SMPL
I1956-01A - 1		15351	4.0550	-	7470A	PPB	24 Feb 2017 10:14:58	U	SMPL
CCV70 - 1	CCV70	18400	4.8617	-	7470A	PPB	24 Feb 2017 10:17:05	U	SMPL
CCB70 - 1	CCB70	-16	-0.0110	-	7470A	PPB	24 Feb 2017 10:19:13	U	SMPL