


245.1

LB86013
INSTRUMENT ID: CY1

Linear

 μ Abs.:

36306



A= 0.0000e+000

B= 2.7439e-004

C= -3.0900e-002

Rho= 0.9999314

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.024	-0.024	24	0.000	24				
0.2	0.200	0.183	-0.017	780	0.0 %	780				
2.5	2.500	2.505	0.005	9243	0.0 %	9243				
5.0	5.000	5.047	0.047	18507	0.0 %	18507				
7.5	7.500	7.558	0.058	27656	0.0 %	27656				
10.0	10.000	9.931	-0.069	36306	0.0 %	36306				

LB86013 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 24 Feb 2017 10:27:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	24	-	0.0000	245.1	PPB	24 Feb 2017 10:30:44	S	Std
0.2 - 1	50.2	780	-	0.2000	245.1	PPB	24 Feb 2017 10:32:52	S	Std
2.5 - 1	52.5	9243	-	2.5000	245.1	PPB	24 Feb 2017 10:35:01	S	Std
5.0 - 1	55	18507	-	5.0000	245.1	PPB	24 Feb 2017 10:37:09	S	Std
7.5 - 1	57.5	27656	-	7.5000	245.1	PPB	24 Feb 2017 10:39:17	S	Std
10.0 - 1	510	36306	-	10.0000	245.1	PPB	24 Feb 2017 10:41:24	S	Std
ICV46 - 1	ICV46	14340	3.9038	-245.1		PPB	24 Feb 2017 10:43:56	U	SMPL
ICB46 - 1	ICB46	-8	-0.0331	-245.1		PPB	24 Feb 2017 10:46:02	U	SMPL
CCV71 - 1	CCV71	19436	5.3021	-245.1		PPB	24 Feb 2017 10:48:09	U	SMPL
CCB71 - 1	CCB71	-11	-0.0339	-245.1		PPB	24 Feb 2017 10:50:15	U	SMPL
CRA - 1	CRA	770	0.1804	-245.1		PPB	24 Feb 2017 10:52:22	U	SMPL
HighStd - 1	HighStd	37125	10.1558	-245.1		PPB	24 Feb 2017 10:54:29	U	SMPL
ChkStd - 1	ChkStd	25301	6.9114	-245.1		PPB	24 Feb 2017 10:56:36	U	SMPL
PB97140BL - 1	PBW	-46	-0.0435	-245.1		PPB	24 Feb 2017 10:58:43	U	SMPL
PB97140BS - 1	LCSW	14731	4.0111	-245.1		PPB	24 Feb 2017 11:00:50	U	SMPL
I1939-01 - 1	021717-PRRT-E-D-M	60	-0.0144	-245.1		PPB	24 Feb 2017 11:02:58	U	SMPL
I1939-02 - 1	021717-PRRT-E-U-M	108	-0.0013	-245.1		PPB	24 Feb 2017 11:05:06	U	SMPL
I1947-01 - 1	022117-PRRT-E-D-M	53	-0.0164	-245.1		PPB	24 Feb 2017 11:07:15	U	SMPL
I1947-02 - 1	022117-PRRT-E-U-M	70	-0.0117	-245.1		PPB	24 Feb 2017 11:09:23	U	SMPL
I1947-03 - 1	022117-DSDC-E-U-M	8	-0.0287	-245.1		PPB	24 Feb 2017 11:11:30	U	SMPL
CCV72 - 1	CCV72	19415	5.2964	-245.1		PPB	24 Feb 2017 11:13:37	U	SMPL
CCB72 - 1	CCB72	-17	-0.0356	-245.1		PPB	24 Feb 2017 11:15:43	U	SMPL
I1947-04 - 1	022117-DSDC-E-D-M	25	-0.0240	-245.1		PPB	24 Feb 2017 11:17:50	U	SMPL
I1969-01 - 1	022217-PRRT-F-D-M	36	-0.0210	-245.1		PPB	24 Feb 2017 11:19:57	U	SMPL
I1969-02 - 1	022217-PRRT-F-U-M	30	-0.0227	-245.1		PPB	24 Feb 2017 11:22:04	U	SMPL
I1970-01 - 1	022317-PRRT-F-D-M	38	-0.0205	-245.1		PPB	24 Feb 2017 11:24:11	U	SMPL
I1970-02 - 1	022317-PRRT-F-U-M	22	-0.0249	-245.1		PPB	24 Feb 2017 11:26:19	U	SMPL
I1970-02DUP - 1	022317-PRRT-F-U-MDUP	39	-0.0202	-245.1		PPB	24 Feb 2017 11:28:26	U	SMPL
I1970-02MS - 1	022317-PRRT-F-U-MMS	16054	4.3741	-245.1		PPB	24 Feb 2017 11:30:35	U	SMPL
I1970-02MSD - 1	022317-PRRT-F-U-MMSD	16394	4.4674	-245.1		PPB	24 Feb 2017 11:32:43	U	SMPL
I1970-02LX5 - 1		-19	-0.0361	-245.1		PPB	24 Feb 2017 11:34:52	U	SMPL
I1970-02A - 1		14811	4.0331	-245.1		PPB	24 Feb 2017 11:36:59	U	SMPL
CCV73 - 1	CCV73	19899	5.4292	-245.1		PPB	24 Feb 2017 11:39:06	U	SMPL
CCB73 - 1	CCB73	-20	-0.0364	-245.1		PPB	24 Feb 2017 11:41:12	U	SMPL