

LB86014



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908-789-8900 Fax: 908-789-892

Analysis Method: 300 Nitrate+Nitrite
Parameter: Nitrate+Nitrite
Run Number: LB86014
Instrument: IC Metrohm

m300.0 - Inorganic Anions - 16

Analytical Summary Report


Analysis Method: 300 Nitrate+Nitrite
Parameter: Nitrate+Nitrite
Run Number: LB86014
Instrument: IC Metrohm

Seq	Lab ID	Client ID	Sample Type	Inst Conc. ppm/ppb	Dil	Analytical Date
1	STD1	STD1	CAL	0.000	1	2/14/17
2	STD2	STD2	CAL	0.759	1	2/14/17
3	STD3	STD3	CAL	1.452	1	2/14/17
4	STD4	STD4	CAL	2.362	1	2/14/17
5	STD5	STD5	CAL	5.279	1	2/14/17
6	STD6	STD6	CAL	10.410	1	2/14/17
7	STD7	STD7	CAL	13.003	1	2/14/17
8	STD8	STD8	CAL	28.061	1	2/14/17
9	ICV	ICV	ICV	13.262	1	2/14/17
10	ICB	ICB	ICB	0.000	1	2/14/17
11	CCV	CCV	CCV	12.665	1	2/23/17
12	CCB	CCB	CCB	0.000	1	2/23/17
13	LB86014BLW	LB86014BLW	MB	0.000	1	2/23/17
14	LB86014BSW	LB86014BSW	LCS	12.964	1	2/23/17
15	I1979-01	2AVE-02	SAM	0.812	1	2/23/17
16	I1979-01D	2AVE-02D	DUP	0.812	1	2/23/17
17	I1979-01S	2AVE-02S	MS	13.308	1	2/23/17
18	CCV	CCV	CCV	12.979	1	2/23/17
19	CCB	CCB	CCB	0.000	1	2/23/17

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/25/17
 Analyst : AK
 Data File : LB86014.MDB

Approved By : 
 Approved Date : 2/25/17
 Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method			Line 2
Nitrate+Nitrite											
STD1	STD1				W	2/14/17					
Nitrate+Nitrite		PASS	0.000		0.000					mg/L	
STD2	STD2				W	2/14/17					
Nitrate+Nitrite		PASS	0.759		0.760					mg/L	
STD3	STD3				W	2/14/17					
Nitrate+Nitrite		PASS	1.452		1.45					mg/L	
STD4	STD4				W	2/14/17					
Nitrate+Nitrite		PASS	2.362		2.36					mg/L	
STD5	STD5				W	2/14/17					
Nitrate+Nitrite		PASS	5.279		5.28					mg/L	
STD6	STD6				W	2/14/17					
Nitrate+Nitrite		PASS	10.410		10.40					mg/L	
STD7	STD7				W	2/14/17					
Nitrate+Nitrite		PASS	13.003		13.00					mg/L	
STD8	STD8				W	2/14/17					
Nitrate+Nitrite		PASS	28.061		28.10					mg/L	
ICV1	ICV1				W	2/14/17					
Nitrate+Nitrite		PASS	13.262		13.30	96.0	90	110		mg/L	
ICB1	ICB1				W	2/14/17					
Nitrate+Nitrite		PASS	0.000		0.000			+/-0.2500		mg/L	
CCV1	CCV1				W	2/23/17					
Nitrate+Nitrite		PASS	12.665		12.70	96.0	90	110		mg/L	
CCB1	CCB1				W	2/23/17					
Nitrate+Nitrite		PASS	0.000		0.000			+/-0.2500		mg/L	
LB86014BLW	LB86014BLW				W	2/23/17					
Nitrate+Nitrite		PASS	0.000		0.000			+/-0.2500		mg/L	
LB86014BSW	LB86014BSW				W	2/23/17					
Nitrate+Nitrite		PASS	12.964		13.00	94.0	90	110		mg/L	
I1979-01	2AVE-02			1	W	2/23/17					
Nitrate+Nitrite		PASS	0.812		0.810					mg/L	
I1979-01D	2AVE-02D			1	W	2/23/17					
Nitrate+Nitrite		PASS	0.812		0.81				0	20	mg/L
I1979-01S	2AVE-02S			1	W	2/23/17					
Nitrate+Nitrite		PASS	13.308		13.30	95.0	80	120		mg/L	
CCV2	CCV2				W	2/23/17					
Nitrate+Nitrite		PASS	12.979		13.00	98.0	90	110		mg/L	
CCB2	CCB2				W	2/23/17					
Nitrate+Nitrite		PASS	0.000		0.000			+/-0.2500		mg/L	