


7471A

LB86016
INSTRUMENT ID: CV1

Linear

 μ Abs.:

47854



A= 0.0000e+000

B= 2.1157e-004

C= 1.1468e-002

Rho= 0.9995658

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.015	0.015	19	0.000	19				
0.2	0.200	0.215	0.015	961	0.0 %	961				
2.5	2.500	2.613	0.113	12296	0.0 %	12296				
5.0	5.000	4.849	-0.151	22864	0.0 %	22864				
7.5	7.500	7.372	-0.128	34790	0.0 %	34790				
10.0	10.000	10.136	0.136	47854	0.0 %	47854				

LB86016 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 24 Feb 2017 11:52:08

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	19	-	0.0000	7471A	PPB	24 Feb 2017 11:55:41	S	Std
0.2 - 1	50.2	961	-	0.2000	7471A	PPB	24 Feb 2017 11:57:58	S	Std
2.5 - 1	52.5	12296	-	2.5000	7471A	PPB	24 Feb 2017 12:00:16	S	Std
5.0 - 1	55	22864	-	5.0000	7471A	PPB	24 Feb 2017 12:02:33	S	Std
7.5 - 1	59.5	34790	-	7.5000	7471A	PPB	24 Feb 2017 12:04:50	S	Std
10.0 - 1	510	47854	-	10.0000	7471A	PPB	24 Feb 2017 12:07:06	S	Std
ICV47 - 1	ICV47	17986	3.8168	-	7471A	PPB	24 Feb 2017 12:09:57	U	SMPL
ICB47 - 1	ICB47	-26	0.0060	-	7471A	PPB	24 Feb 2017 12:12:12	U	SMPL
CCV74 - 1	CCV74	22950	4.8670	-	7471A	PPB	24 Feb 2017 12:14:28	U	SMPL
CCB74 - 1	CCB74	-17	0.0079	-	7471A	PPB	24 Feb 2017 12:16:43	U	SMPL
CRA - 1	CRA	977	0.2182	-	7471A	PPB	24 Feb 2017 12:18:59	U	SMPL
HighStd - 1	HighStd	48312	10.2329	-	7471A	PPB	24 Feb 2017 12:21:15	U	SMPL
ChkStd - 1	ChkStd	33430	7.0843	-	7471A	PPB	24 Feb 2017 12:23:31	U	SMPL
PB97147BL - 1	PBS	-15	0.0083	-	7471A	PPB	24 Feb 2017 12:25:48	U	SMPL
PB97147BS - 1	LCSS	16346	3.4698	-	7471A	PPB	24 Feb 2017 12:28:05	U	SMPL
I1948-05 - 1	TP23-24-COMP17	10296	2.1898	-	7471A	PPB	24 Feb 2017 12:30:21	U	SMPL
I1952-01 - 1	SPARTA-FINE-SAND	527	0.1230	-	7471A	PPB	24 Feb 2017 12:32:39	U	SMPL
I1952-03 - 1	SPARTA-TOP-SOIL	2306	0.4993	-	7471A	PPB	24 Feb 2017 12:34:56	U	SMPL
I1954-01 - 1	CLP-S-04(0-5)	2889	0.6227	-	7471A	PPB	24 Feb 2017 12:37:14	U	SMPL
I1954-02 - 1	CLP-S-04(5-10)	317	0.0785	-	7471A	PPB	24 Feb 2017 12:39:30	U	SMPL
CCV75 - 1	CCV75	23180	4.9157	-	7471A	PPB	24 Feb 2017 12:41:46	U	SMPL
CCB75 - 1	CCB75	-15	0.0083	-	7471A	PPB	24 Feb 2017 12:44:02	U	SMPL
I1954-02DUP - 1	CLP-S-04(5-10)DUP	344	0.0842	-	7471A	PPB	24 Feb 2017 12:49:00	U	SMPL
I1954-02MS - 1	CLP-S-04(5-10)MS	16824	3.5709	-	7471A	PPB	24 Feb 2017 12:51:15	U	SMPL
I1954-02MSD - 1	CLP-S-04(5-10)MSD	16449	3.4916	-	7471A	PPB	24 Feb 2017 12:53:32	U	SMPL
I1954-03 - 1	CLP-S-04(10-15)	137	0.0405	-	7471A	PPB	24 Feb 2017 12:55:48	U	SMPL
I1954-05 - 1	CLP-S-06(0-5)	2559	0.5529	-	7471A	PPB	24 Feb 2017 12:58:05	U	SMPL
I1954-06 - 1	CLP-S-06(5-10)	5660	1.2090	-	7471A	PPB	24 Feb 2017 13:00:22	U	SMPL
I1954-07 - 1	CLP-S-06(10-15)	2870	0.6187	-	7471A	PPB	24 Feb 2017 13:02:39	U	SMPL
I1954-09 - 1	CLP-B-01(0-5)	2495	0.5393	-	7471A	PPB	24 Feb 2017 13:04:57	U	SMPL
I1954-10 - 1	CLP-B-01(5-10)	7449	1.5875	-	7471A	PPB	24 Feb 2017 13:07:15	U	SMPL
I1954-11 - 1	CLP-B-01(10-15)	151	0.0434	-	7471A	PPB	24 Feb 2017 13:09:31	U	SMPL
CCV76 - 1	CCV76	23408	4.9639	-	7471A	PPB	24 Feb 2017 13:11:46	U	SMPL
CCB76 - 1	CCB76	-11	0.0091	-	7471A	PPB	24 Feb 2017 13:14:03	U	SMPL
I1954-13 - 1	CLP-S-03(0-5)	11676	2.4818	-	7471A	PPB	24 Feb 2017 13:16:19	U	SMPL
I1954-14 - 1	CLP-S-03(0-5)-D	5458	1.1662	-	7471A	PPB	24 Feb 2017 13:18:36	U	SMPL
I1954-15 - 1	CLP-S-03(5-10)	1161	0.2571	-	7471A	PPB	24 Feb 2017 13:20:52	U	SMPL
I1954-16 - 1	CLP-S-03(10-15)	276	0.0699	-	7471A	PPB	24 Feb 2017 13:23:09	U	SMPL
I1955-03 - 1	B1-3	226678	47.9698	-	7471A	PPB	24 Feb 2017 13:25:25	U	SMPL
I1955-07 - 1	B1-4	192264	40.6889	-	7471A	PPB	24 Feb 2017 13:27:42	U	SMPL
I1966-02 - 1	33B-9-11	260	0.0665	-	7471A	PPB	24 Feb 2017 13:30:02	U	SMPL
PB97148BL - 1	PBS	-8	0.0098	-	7471A	PPB	24 Feb 2017 13:32:21	U	SMPL
PB97148BS - 1	LCSS	16473	3.4967	-	7471A	PPB	24 Feb 2017 13:34:39	U	SMPL
I1957-01 - 1	SRI-SB-3(8-10)20170221	751	0.1704	-	7471A	PPB	24 Feb 2017 13:36:56	U	SMPL
CCV77 - 1	CCV77	23090	4.8966	-	7471A	PPB	24 Feb 2017 13:39:12	U	SMPL
CCB77 - 1	CCB77	-11	0.0091	-	7471A	PPB	24 Feb 2017 13:41:29	U	SMPL
I1957-02 - 1	SRI-SB-3(12-14)20170221	106	0.0339	-	7471A	PPB	24 Feb 2017 13:43:45	U	SMPL
I1957-03 - 1	SRI-SB-4(8-10)20170221	578	0.1338	-	7471A	PPB	24 Feb 2017 13:46:01	U	SMPL
I1957-04 - 1	SRI-SB-4(10-11)20170221	119	0.0366	-	7471A	PPB	24 Feb 2017 13:48:18	U	SMPL
I1957-05 - 1	SRI-SB-6(8-10)20170222	3075	0.6620	-	7471A	PPB	24 Feb 2017 13:50:34	U	SMPL
I1957-05DUP - 1	SRI-SB-6(8-10)20170222DUP	3043	0.6553	-	7471A	PPB	24 Feb 2017 13:52:51	U	SMPL
I1957-06 - 1	I1957-05MS	19351	4.1056	-	7471A	PPB	24 Feb 2017 13:55:08	U	SMPL
I1957-07 - 1	I1957-05MSD	19049	4.0417	-	7471A	PPB	24 Feb 2017 13:57:25	U	SMPL
I1957-08 - 1	SRI-SB-6(12-14)20170222	631	0.1450	-	7471A	PPB	24 Feb 2017 13:59:43	U	SMPL
I1957-09 - 1	SRI-SB-5(8-10)20170222	909	0.2038	-	7471A	PPB	24 Feb 2017 14:02:01	U	SMPL
I1957-10 - 1	SRI-SB-5(13-15)20170222	418	0.0999	-	7471A	PPB	24 Feb 2017 14:04:17	U	SMPL
CCV78 - 1	CCV78	22748	4.8243	-	7471A	PPB	24 Feb 2017 14:06:33	U	SMPL
CCB78 - 1	CCB78	-1	0.0113	-	7471A	PPB	24 Feb 2017 14:08:50	U	SMPL
I1957-11 - 1	SRI-SB-7(8-9)20170222	155	0.0443	-	7471A	PPB	24 Feb 2017 14:11:06	U	SMPL
I1957-14 - 1	SRI-SB-X(8-10)20170222	1689	0.3688	-	7471A	PPB	24 Feb 2017 14:13:22	U	SMPL
I1957-15 - 1	SRI-SB-5(16-18)20170222	1340	0.2950	-	7471A	PPB	24 Feb 2017 14:15:39	U	SMPL
I1954-02LX5 - 1		72	0.0267	-	7471A	PPB	24 Feb 2017 14:17:55	U	SMPL
I1954-02A - 1		17087	3.6266	-	7471A	PPB	24 Feb 2017 14:20:12	U	SMPL
I1957-05LX5 - 1		601	0.1386	-	7471A	PPB	24 Feb 2017 14:22:29	U	SMPL
I1957-05A - 1		19530	4.1434	-	7471A	PPB	24 Feb 2017 14:24:46	U	SMPL
I1955-03DLX10 - 1		24173	5.1258	-	7471A	PPB	24 Feb 2017 14:27:04	U	SMPL
I1955-07DLX10 - 1		20595	4.3688	-	7471A	PPB	24 Feb 2017 14:29:22	U	SMPL
CCV79 - 1	CCV79	22899	4.8562	-	7471A	PPB	24 Feb 2017 14:31:38	U	SMPL
CCB79 - 1	CCB79	-4	0.0106	-	7471A	PPB	24 Feb 2017 14:33:55	U	SMPL