



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86020

Temp.: 18.6 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/24/2017	12:35
2	CAL2	1	Water			7.00	02/24/2017	12:36
3	CAL3	1	Water			10.03	02/24/2017	12:37
4	ICV	1	Water			7.01	02/24/2017	12:38
5	CCV1	1	Water			2.02	02/24/2017	12:40
6	I1983-01	1	Water			8.79	02/24/2017	12:41
7	I1983-01DUP	1	Water			8.80	02/24/2017	12:42
8	I1985-01	1	Water			7.27	02/24/2017	12:43
9	CCV2	1	Water			12.01	02/24/2017	12:45

1386020

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I1983

WorkList ID : 95941

Date : 2/24/2017 12:21:49 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1983-01	pH	Cool 4 deg C	VIA01	I53	EFF-170223-01	02/23/2017	SM 4500-PH B
	Water	I1985-01	pH	Cool 4 deg C	VIA01	I53	INF-170223-01A	02/23/2017	SM 4500-PH B

Date/Time 02-24-17 12:21 PM
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 02-24-17 21:00 PM
Received by: [Signature]
Relinquished by: [Signature]