



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: JIGNESHDate: 02/24/2017Run Number: LB86023BalanceID: WC SC-3DATE & TIME IN: 02/24/2017 14:20DATE & TIME OUT: 02/24/2017 18:30OVEN TEMP: 105 °COVEN TEMP: 104 °C

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Weight (g)	Result mg/L
1	LB86023BL	LB86023BL	1.4384	1.4384	100	1.4384	1.4384	1.4384	0	0
2	LB86023BS	LB86023BS	1.3687	1.3687	100	1.4234	1.4234	1.4234	0.0547	547
3	I1983-01	EFF-170223-01	1.4099	1.4099	800	1.412	1.412	1.412	0.0021	2.6
4	I1985-01	INF-170223-01A	1.4216	1.4216	300	1.4318	1.4318	1.4318	0.0102	34
5	I1985-01DUP	INF-170223-01ADUP	1.4155	1.4155	300	1.4261	1.4261	1.4261	0.0106	35.3

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

Date : 2/24/2017 12:54:09 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	11983-01-E	TSS	Cool 4 deg C	VIAS01	I53	EFF-170223-01	02/23/2017	SM2540 D
	Water	11985-01-E	TSS	Cool 4 deg C	VIAS01	I53	INF-170223-01A	02/23/2017	SM2540 D

Date/Time 02-26-17 12:54 PM
Received by: J
Relinquished by:

Date/Time 02-23-17 3:50 PM
Received by: AD
Relinquished by: SP