



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB86025

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	02/24/2017	13:40
2	I1956-02		1	>100.0	>212.0	02/24/2017	14:15
3	I1956-02DUP		1	>100.0	>212.0	02/24/2017	14:45
4	I1979-01		1	>100.0	>212.0	02/24/2017	15:20

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 86025

Worklist Name : FP22417

Worklist ID : 95946

Date : 2/24/2017 1:00 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/27/2017	Water	I1956-02 B	Flash Point	Cool 4 deg C	PSEG01	F61	FRAC-TANK-A2622	02/22/2017	1010A
	Water	I1979-01 T	Flash Point	Cool 4 deg C	DIST01	I61	2AVE-02	02/23/2017	1010A

Date/Time

2/24/17 1:10 PM

Received by:

KS

Relinquished by:

CP

Date/Time

2/24/17 15:30

Received by:

CP

Relinquished by:

KS