


7471A

LB 86042
INSTRUMENT ID: CV1

Linear

 μ Abs.:

47345



A= 0.0000e+000

B= 2.1103e-004

C= 2.4750e-002

Rho= 0.9999597

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.025	0.025	1	0.000	1				
0.2	0.200	0.212	0.012	887	0.0 %	887				
2.5	2.500	2.438	-0.062	11434	0.0 %	11434				
5.0	5.000	5.033	0.033	23735	0.0 %	23735				
7.5	7.500	7.476	-0.024	35311	0.0 %	35311				
10.0	10.000	10.016	0.016	47345	0.0 %	47345				

LB86042 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 27 Feb 2017 12:42:13

Sample ID	Extended ID	µ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	1	-	0.00007471A	PPB	27 Feb 2017 12:45:18	S	Std
0.2 - 1	50.2	887	-	0.20007471A	PPB	27 Feb 2017 12:47:35	S	Std
2.5 - 1	52.5	11434	-	2.50007471A	PPB	27 Feb 2017 12:49:53	S	Std
5.0 - 1	55	23735	-	5.00007471A	PPB	27 Feb 2017 12:52:10	S	Std
7.5 - 1	57.5	35311	-	7.50007471A	PPB	27 Feb 2017 12:54:28	S	Std
10.0 - 1	50	47345	-	10.00007471A	PPB	27 Feb 2017 12:56:44	S	Std
ICV49 - 1	ICV49	18610	3.9519	-7471A	PPB	27 Feb 2017 12:59:25	U	SMPL
ICB49 - 1	ICB49	-76	0.0087	-7471A	PPB	27 Feb 2017 13:01:40	U	SMPL
CCV83 - 1	CCV83	22289	4.7283	-7471A	PPB	27 Feb 2017 13:03:56	U	SMPL
CCB83 - 1	CCB83	-73	0.0093	-7471A	PPB	27 Feb 2017 13:06:12	U	SMPL
CRA - 1	CRA	894	0.2134	-7471A	PPB	27 Feb 2017 13:08:28	U	SMPL
HighStd - 1	HighStd	47124	9.9691	-7471A	PPB	27 Feb 2017 13:10:43	U	SMPL
ChkStd - 1	ChkStd	33401	7.0732	-7471A	PPB	27 Feb 2017 13:12:59	U	SMPL
PB97177BL - 1	PBS	-62	0.0117	-7471A	PPB	27 Feb 2017 13:15:16	U	SMPL
PB97177BS - 1	LCSS	17944	3.8114	-7471A	PPB	27 Feb 2017 13:17:33	U	SMPL
I1972-01 - 1	E2B15-BASE-1	30425	6.4452	-7471A	PPB	27 Feb 2017 13:19:50	U	SMPL
I1972-02 - 1	E2B15-BASE-2	224045	47.3040	-7471A	PPB	27 Feb 2017 13:22:07	U	SMPL
I1972-03 - 1	E2B15-BASE-3	62809	13.2791	-7471A	PPB	27 Feb 2017 13:24:24	U	SMPL
I1972-04 - 1	E2B21-BASE-1	107766	22.7662	-7471A	PPB	27 Feb 2017 13:26:44	U	SMPL
I1972-05 - 1	E2B21-BASE-2	56256	11.8962	-7471A	PPB	27 Feb 2017 13:29:01	U	SMPL
CCV84 - 1	CCV84	22998	4.8779	-7471A	PPB	27 Feb 2017 13:31:18	U	SMPL
CCB84 - 1	CCB84	-85	0.0068	-7471A	PPB	27 Feb 2017 13:33:35	U	SMPL
I1972-06 - 1	E2B21-BASE-3	469118	HIGH	-7471A	PPB	27 Feb 2017 13:35:51	X	
I1972-07 - 1	E2B21-WESTSW-1	45837	9.6975	-7471A	PPB	27 Feb 2017 13:37:40	U	SMPL
I1972-08 - 1	E2B21-WESTSW-2	103658	21.8993	-7471A	PPB	27 Feb 2017 13:39:59	U	SMPL
I1972-09 - 1	E2-B1-BASE-1	323801	68.3551	-7471A	PPB	27 Feb 2017 13:42:17	U	SMPL
I1972-10 - 1	E2-B1-BASE-2	729601	HIGH	-7471A	PPB	27 Feb 2017 13:44:35	X	
I1972-11 - 1	E2-B1-BASE-3	24765	5.2508	-7471A	PPB	27 Feb 2017 13:46:28	U	SMPL
I1972-12 - 1	E2-B1-SW-5	489243	HIGH	-7471A	PPB	27 Feb 2017 13:48:49	X	
I1972-13 - 1	E2-B1-SW-4	340952	HIGH	-7471A	PPB	27 Feb 2017 13:50:40	X	
I1972-14 - 1	E2-B1-SW-3	224436	47.3865	-7471A	PPB	27 Feb 2017 13:52:36	U	SMPL
I1972-15 - 1	E2-B1-SW-2	113193	23.9114	-7471A	PPB	27 Feb 2017 13:54:54	U	SMPL
CCV85 - 1	CCV85	22870	4.8509	-7471A	PPB	27 Feb 2017 13:57:13	U	SMPL
CCB85 - 1	CCB85	-88	0.0062	-7471A	PPB	27 Feb 2017 13:59:31	U	SMPL
I1972-16 - 1	E2-B1-SW-1	99502	21.0222	-7471A	PPB	27 Feb 2017 14:01:47	U	SMPL
I1972-17 - 1	E2HH8.5-BASE	721390	HIGH	-7471A	PPB	27 Feb 2017 14:04:03	X	
I1972-18 - 1	E2-B1-BASE-2-02	135971	28.7181	-7471A	PPB	27 Feb 2017 14:05:54	U	SMPL
I1972-19 - 1	E2-HH5-SW-1	194953	41.1649	-7471A	PPB	27 Feb 2017 14:08:13	U	SMPL
I1972-20 - 1	E2-HH5-SW-2	413503	HIGH	-7471A	PPB	27 Feb 2017 14:10:32	X	
I1972-21 - 1	E2-Z5-BASE	41759	8.8370	-7471A	PPB	27 Feb 2017 14:12:25	U	SMPL
I1972-22 - 1	E2-Z6-BASE	1241	0.2866	-7471A	PPB	27 Feb 2017 14:14:44	U	SMPL
I1972-22DUP - 1	E2-Z6-BASEDUP	1245	0.2875	-7471A	PPB	27 Feb 2017 14:17:01	U	SMPL
I1972-22MS - 1	E2-Z6-BASEMS	16213	3.4461	-7471A	PPB	27 Feb 2017 14:19:19	U	SMPL
I1972-22MSD - 1	E2-Z6-BASEMSD	15867	3.3731	-7471A	PPB	27 Feb 2017 14:21:36	U	SMPL
CCV86 - 1	CCV86	23109	4.9013	-7471A	PPB	27 Feb 2017 14:23:52	U	SMPL
CCB86 - 1	CCB86	-51	0.0140	-7471A	PPB	27 Feb 2017 14:26:08	U	SMPL
I1972-23 - 1	E2-Y5N-SW2	2841	0.6243	-7471A	PPB	27 Feb 2017 14:28:25	U	SMPL
PB97178BL - 1	PBS	19	0.0288	-7471A	PPB	27 Feb 2017 14:30:41	U	SMPL
PB97178BS - 1	LCSS	16470	3.5003	-7471A	PPB	27 Feb 2017 14:32:58	U	SMPL
I1973-01 - 1	E2-Y6-BASE	1472	0.3354	-7471A	PPB	27 Feb 2017 14:35:14	U	SMPL
I1973-02 - 1	E2-Y5-BASE	1293	0.2976	-7471A	PPB	27 Feb 2017 14:37:31	U	SMPL
I1973-03 - 1	E2-Z5W-SW	68527	14.4857	-7471A	PPB	27 Feb 2017 14:39:47	U	SMPL
I1973-04 - 1	E2-Y5S-SW	75473	15.9515	-7471A	PPB	27 Feb 2017 14:42:05	U	SMPL
I1973-05 - 1	E2-Y5N-SW	23113	4.9022	-7471A	PPB	27 Feb 2017 14:44:23	U	SMPL
I1973-06 - 1	E2-Y6S-SW	260123	54.9174	-7471A	PPB	27 Feb 2017 14:46:42	U	SMPL
I1973-07 - 1	E2-Y6N-SW	897	0.2140	-7471A	PPB	27 Feb 2017 14:48:58	U	SMPL
CCV87 - 1	CCV87	23291	4.9398	-7471A	PPB	27 Feb 2017 14:51:17	U	SMPL
CCB87 - 1	CCB87	-62	0.0117	-7471A	PPB	27 Feb 2017 14:53:33	U	SMPL
I1973-08 - 1	E2-Y6E-SW	895	0.2136	-7471A	PPB	27 Feb 2017 14:55:50	U	SMPL
I1973-09 - 1	E2-FF9-BASE	49443	10.4585	-7471A	PPB	27 Feb 2017 14:58:06	U	SMPL
I1973-10 - 1	E2-GG9-BASE	82553	17.4456	-7471A	PPB	27 Feb 2017 15:00:23	U	SMPL
I1973-11 - 1	E2-B1-B2-BASE-1	66137	13.9814	-7471A	PPB	27 Feb 2017 15:02:39	U	SMPL
I1973-12 - 1	E2-B1-B2-BASE-2	34876	7.3845	-7471A	PPB	27 Feb 2017 15:04:56	U	SMPL
I1973-13 - 1	E2-B1-B2-BASE-3	143355	30.2763	-7471A	PPB	27 Feb 2017 15:07:14	U	SMPL
I1973-14 - 1	E2-B1-B2-BASE-4	861219	HIGH	-7471A	PPB	27 Feb 2017 15:09:31	X	
I1973-15 - 1	E2-B1-B2-BASE-5	29246	6.1964	-7471A	PPB	27 Feb 2017 15:16:26	U	SMPL
I1973-15DUP - 1	E2-B1-B2-BASE-5DUP	29101	6.1658	-7471A	PPB	27 Feb 2017 15:18:44	U	SMPL
I1973-15MS - 1	E2-B1-B2-BASE-5MS	45624	9.6526	-7471A	PPB	27 Feb 2017 15:25:18	U	SMPL
CCV88 - 1	CCV88	23280	4.9374	-7471A	PPB	27 Feb 2017 15:27:34	U	SMPL
CCB88 - 1	CCB88	-51	0.0140	-7471A	PPB	27 Feb 2017 15:29:50	U	SMPL
I1973-15MSD - 1	E2-B1-B2-BASE-5MSD	45474	9.6209	-7471A	PPB	27 Feb 2017 15:32:07	U	SMPL
I1973-16 - 1	E2-B1-B2-BASE-6	27404	5.8077	-7471A	PPB	27 Feb 2017 15:34:24	U	SMPL
I1973-17 - 1	E2-B1-B2-BASE-7	105153	22.2147	-7471A	PPB	27 Feb 2017 15:36:40	U	SMPL
I1973-18 - 1	E2-B1-B2-SOUTHSW-1	952626	HIGH	-7471A	PPB	27 Feb 2017 15:38:57	X	
I1973-19 - 1	E2-B1-B2-SOUTHSW-2	23893	5.0668	-7471A	PPB	27 Feb 2017 15:40:47	U	SMPL
PB97179BL - 1	PBS	-157	-0.0084	-7471A	PPB	27 Feb 2017 15:43:19	U	SMPL
PB97179BS - 1	LCSS	20325	4.3138	-7471A	PPB	27 Feb 2017 15:45:36	U	SMPL
I1989-01 - 1	FK-BPM-01-CC9-CC10-CC11-CC12	396	0.1083	-7471A	PPB	27 Feb 2017 15:47:53	U	SMPL
I1989-01DUP - 1	FK-BPM-01-CC9-CC10-CC11-CC12DUP	421	0.1136	-7471A	PPB	27 Feb 2017 15:50:10	U	SMPL
I1989-01MS - 1	FK-BPM-01-CC9-CC10-CC11-CC12MS	15920	3.3843	-7471A	PPB	27 Feb 2017 15:52:27	U	SMPL
CV89 - 1	CCV89	23571	4.9988	-7471A	PPB	27 Feb 2017 15:54:44	U	SMPL
CB89 - 1	CCB89	-47	0.0148	-7471A	PPB	27 Feb 2017 15:57:01	U	SMPL
I1989-01MSD - 1	FK-BPM-01-CC9-CC10-CC11-CC12MSD	16286	3.4615	-7471A	PPB	27 Feb 2017 15:59:18	U	SMPL
I1990-01 - 1	FK-BPM-01-BB13-BB14-CC13-CC14	338	0.0961	-7471A	PPB	27 Feb 2017 16:01:35	U	SMPL
I1991-01 - 1	FK-BPM-01-DD13-DD14-EE13-EE14	489	0.1279	-7471A	PPB	27 Feb 2017 16:03:52	U	SMPL
I1992-01 - 1	FK-GF-02-003-A	2154	0.4793	-7471A	PPB	27 Feb 2017 16:06:09	U	SMPL
I1992-03 - 1	FK-GF-02-003-B	3516	0.7667	-7471A	PPB	27 Feb 2017 16:08:26	U	SMPL
I1995-01 - 1	U-1	1373	0.3145	-7471A	PPB	27 Feb 2017 16:10:43	U	SMPL
I1995-03 - 1	U-2	1115	0.2600	-7471A	PPB	27 Feb 2017 16:13:01	U	SMPL
CV90 - 1	CCV90	23417	4.9663	-7471A	PPB	27 Feb 2017 16:15:18	U	SMPL
CB90 - 1	CCB90	-37	0.0169	-7471A	PPB	27 Feb 2017 16:17:36	U	SMPL
I1972-22LX5 - 1		288	0.0855	-7471A	PPB	27 Feb 2017 16:19:53	U	SMPL
I1972-22A - 1		17027	3.6178	-7471A	PPB	27 Feb 2017 16:22:10	U	SMPL

LB86042 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 27 Feb 2017 12:42:13				
Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
I1973-15LX5 - 1	CCV91 CCB91	5730	1.2339	-7471A	PPB	27 Feb 2017 16:24:26	U	SMPL
I1973-15A - 1		45112	9.5445	-7471A	PPB	27 Feb 2017 16:26:43	U	SMPL
I1989-01LX5 - 1		23	0.0296	-7471A	PPB	27 Feb 2017 16:29:00	U	SMPL
I1989-01A - 1		17105	3.6343	-7471A	PPB	27 Feb 2017 16:31:17	U	SMPL
I1972-06DLX10 - 1		74213	15.6856	-7471A	PPB	27 Feb 2017 16:33:34	U	SMPL
I1972-10DLX10 - 1		136144	28.7546	-7471A	PPB	27 Feb 2017 16:35:51	U	SMPL
I1972-12DLX10 - 1		74143	15.6708	-7471A	PPB	27 Feb 2017 16:38:09	U	SMPL
I1972-13DLX10 - 1		45094	9.5407	-7471A	PPB	27 Feb 2017 16:40:28	U	SMPL
CCV91 - 1		23459	4.9752	-7471A	PPB	27 Feb 2017 16:42:47	U	SMPL
CCB91 - 1		-50	0.0142	-7471A	PPB	27 Feb 2017 16:45:04	U	SMPL
I1972-17DLX10 - 1		139807	29.5276	-7471A	PPB	27 Feb 2017 16:47:22	U	SMPL
I1972-20DLX10 - 1		59865	12.6578	-7471A	PPB	27 Feb 2017 16:49:38	U	SMPL
I1973-14DLX10 - 1		189774	40.0720	-7471A	PPB	27 Feb 2017 16:51:57	U	SMPL
I1973-18DLX10 - 1		255416	53.9241	-7471A	PPB	27 Feb 2017 16:57:43	U	SMPL
I1972-02DLX10 - 1		23303	4.9423	-7471A	PPB	27 Feb 2017 16:59:59	U	SMPL
I1972-03DLX5 - 1		12516	2.6659	-7471A	PPB	27 Feb 2017 17:02:18	U	SMPL
I1972-04DLX5 - 1		21950	4.6568	-7471A	PPB	27 Feb 2017 17:04:35	U	SMPL
I1972-05DLX5 - 1		11374	2.4250	-7471A	PPB	27 Feb 2017 17:06:52	U	SMPL
I1972-08DLX5 - 1		20500	4.3508	-7471A	PPB	27 Feb 2017 17:09:09	U	SMPL
I1972-09DLX10 - 1		34959	7.4020	-7471A	PPB	27 Feb 2017 17:11:26	U	SMPL
CCV92 - 1	CCV92 CCB92	23788	5.0446	-7471A	PPB	27 Feb 2017 17:13:44	U	SMPL
CCB92 - 1		-42	0.0159	-7471A	PPB	27 Feb 2017 17:16:02	U	SMPL
I1972-14DLX10 - 1		23230	4.9269	-7471A	PPB	27 Feb 2017 17:18:19	U	SMPL
I1972-15DLX5 - 1		22786	4.8332	-7471A	PPB	27 Feb 2017 17:20:36	U	SMPL
I1972-16DLX5 - 1		19819	4.2071	-7471A	PPB	27 Feb 2017 17:22:53	U	SMPL
I1972-18DLX5 - 1		28202	5.9761	-7471A	PPB	27 Feb 2017 17:25:11	U	SMPL
I1972-19DLX10 - 1		20761	4.4059	-7471A	PPB	27 Feb 2017 17:27:28	U	SMPL
I1973-03DLX5 - 1		13542	2.8825	-7471A	PPB	27 Feb 2017 17:29:45	U	SMPL
I1973-04DLX5 - 1		15105	3.2123	-7471A	PPB	27 Feb 2017 17:32:03	U	SMPL
I1973-06DLX10 - 1		27507	5.8294	-7471A	PPB	27 Feb 2017 17:34:20	U	SMPL
I1973-09DLX5 - 1		9811	2.0951	-7471A	PPB	27 Feb 2017 17:36:37	U	SMPL
I1973-10DLX5 - 1		17208	3.6561	-7471A	PPB	27 Feb 2017 17:38:55	U	SMPL
CCV93 - 1		23468	4.9771	-7471A	PPB	27 Feb 2017 17:41:12	U	SMPL
CCB93 - 1		-39	0.0165	-7471A	PPB	27 Feb 2017 17:43:30	U	SMPL
I1973-11DLX5 - 1		12908	2.7487	-7471A	PPB	27 Feb 2017 17:45:47	U	SMPL
I1973-13DLX5 - 1		29453	6.2401	-7471A	PPB	27 Feb 2017 17:48:05	U	SMPL
I1973-17DLX5 - 1		20808	4.4158	-7471A	PPB	27 Feb 2017 17:50:22	U	SMPL
I1972-06DL2X50 - 1		14380	3.0593	-7471A	PPB	27 Feb 2017 17:52:40	U	SMPL
I1972-10DL2X50 - 1		27600	5.8491	-7471A	PPB	27 Feb 2017 17:54:58	U	SMPL
I1972-12DL2X50 - 1		14932	3.1758	-7471A	PPB	27 Feb 2017 17:57:15	U	SMPL
I1972-17DL2X50 - 1		27699	5.8700	-7471A	PPB	27 Feb 2017 17:59:33	U	SMPL
I1972-20DL2X50 - 1		11480	2.4473	-7471A	PPB	27 Feb 2017 18:01:51	U	SMPL
I1973-14DL2X50 - 1		38647	8.1803	-7471A	PPB	27 Feb 2017 18:04:08	U	SMPL
I1973-18DL2X50 - 1		55659	11.7702	-7471A	PPB	27 Feb 2017 18:06:26	U	SMPL
CCV94 - 1	CCV94 CCB94	23297	4.9410	-7471A	PPB	27 Feb 2017 18:08:43	U	SMPL
CCB94 - 1		-66	0.0108	-7471A	PPB	27 Feb 2017 18:11:01	U	SMPL