



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86046

Temp.: 20.1 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/27/2017	13:25
2	CAL2	1	Water			7.01	02/27/2017	13:26
3	CAL3	1	Water			10.03	02/27/2017	13:27
4	ICV	1	Water			7.02	02/27/2017	13:30
5	CCV1	1	Water			2.02	02/27/2017	13:31
6	I2011-02	1	Water			8.21	02/27/2017	13:32
7	I2011-02DUP	1	Water			8.22	02/27/2017	13:33
8	CCV2	1	Water			12.01	02/27/2017	13:34

WORKLIST(Hardcopy Internal Chain)

2B 86046

WorkList Name : CORROSIVITY I2011

WorkList ID : 95991

Date : 2/27/2017 1:03:34 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/27/2017	Water	I2011-02 -A	Corrosivity	Cool 4 deg C	PSEG01	F62	FRAC-TANK-A155	02/27/2017	9040C

Date/Time 02-27-13 11:04 PM
 Received by: JP
 Relinquished by: CP

Date/Time 02-27-13 2:13:50 PM
 Received by: CP
 Relinquished by: JP