

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
=====	=====	=====	=====	=====	=====
CCV	10.9373			TOC	
CCB	0.3638			TOC	
LB86061BLW	0.1609	0.0697	43.31	TOC	
LB86061BSW.....	9.4703...	0.4086..	4.31...	TOC	..
I2004-01	5.0442	1.8722	37.12	TOC	
I2004-02	1.3872	0.1192	8.59	TOC	
I2004-03.....	2.5017...	0.3900..	15.59...	TOC	..
I2004-04	3.8329	0.2009	5.24	TOC	
I2004-04S	13.4193	0.2942	2.19	TOC	
I2004-04SD.....	11.7398...	1.0510..	8.95...	TOC	..
CCV	10.2024			TOC	
CCB	0.3971			TOC	

Sample Results Report Print Date/Time: 2017/02/28 14:31:14

Method ID	Sample Type	Vial	Timestamp	Message
TOC 0 - 20 ppmC	Sample	2	2017/02/28 08:35	
TOC 0 - 20 ppmC	Sample	3	2017/02/28 08:49	
TOC 0 - 20 ppmC	Sample	4	2017/02/28 09:13	Low Sample Detected
TOC 0 - 20 ppmC	...Sample	.. 5..	2017/02/28 10:07	..
TOC 0 - 20 ppmC	Sample	6	2017/02/28 10:30	
TOC 0 - 20 ppmC	Sample	7	2017/02/28 10:52	
TOC 0 - 20 ppmC	...Sample	.. 8..	2017/02/28 11:15	..
TOC 0 - 20 ppmC	Sample	9	2017/02/28 11:37	
TOC 0 - 20 ppmC	Sample	10	2017/02/28 12:03	
TOC 0 - 20 ppmC	...Sample	.. 11..	2017/02/28 12:27	..
TOC 0 - 20 ppmC	Sample	2	2017/02/28 14:05	
TOC 0 - 20 ppmC	Sample	3	2017/02/28 14:18	

AK 21117

Calibration Report Print Date/Time: 2017/01/31 17:01:59

Cal. Curve ID: TOC WATER 0-20PPM
 Created: 2017/01/30 12:41
 Calibration Factor (m): 5.417e+04
 Y Intercept (b): 9373
 r-squared: 0.99919

Standard ID	Y	X Expected	Measured	Message	Date & Time
	Raw Data	ug C	ug C		
0.0PPM	7751	0.000	-0.030		2017/01/30 10:36
0.5PPM	21494	0.250	0.224		2017/01/30 10:49
1.0PPM	34457	0.500	0.463		2017/01/30 11:42
5.0PPM	142741	2.500	2.462		2017/01/30 11:55
10.0PPM	292332	5.000	5.224		2017/01/30 12:08
20.0PPM	545609	10.000	9.899		2017/01/30 12:22
2.0PPM	63973	1.000	1.008		2017/01/30 12:40

AK 21117

Sample Results Report Print Date/Time: 2017/01/31 17:02:48

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
0.0PPM	7751			TOC	
0.5PPM	21494			TOC	
1.0PPM	34457			TOC	
5.0PPM.....	142741...	..	...	TOC	..
10.0PPM	292332			TOC	
20.0PPM	545609			TOC	
2.0PPM.....	63973...	..	...	TOC	..
ICV	9.6706			TOC	
ICB	0.2636			TOC	

AK 2/1/17

Method ID	Sample Type	Vial	Timestamp	Message
TOC 0 - 20 ppmC	TOC Standard	1	2017/01/30 10:36	
TOC 0 - 20 ppmC	TOC Standard	2	2017/01/30 10:49	
TOC 0 - 20 ppmC	TOC Standard	3	2017/01/30 11:42	
TOC 0 - 20 ppmC	...TOC Standard	.. 5..	2017/01/30 11:55	..
TOC 0 - 20 ppmC	TOC Standard	6	2017/01/30 12:08	
TOC 0 - 20 ppmC	TOC Standard	7	2017/01/30 12:22	
TOC 0 - 20 ppmC	...TOC Standard	.. 4..	2017/01/30 12:40	..
TOC 0 - 20 ppmC	Sample	8	2017/01/30 12:55	
TOC 0 - 20 ppmC	Sample	9	2017/01/30 13:08	