



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86062

Temp.: 24.4 °C

BalanceID: WC SC-4

SLOPE.: 99.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/27/2017	16:20
2	CAL2	1	Water			7.01	02/27/2017	16:21
3	CAL3	1	Water			10.02	02/27/2017	16:22
4	ICV	1	Water			7.02	02/27/2017	16:23
5	CCV1	1	Water			2.01	02/27/2017	16:25
6	I2014-01	1	Solid	20.02	20	6.12	02/27/2017	16:26
7	I2014-01DUP	1	Solid	20.04	20	6.13	02/27/2017	16:27
8	CCV2	1	Water			12.02	02/27/2017	16:29

1386062

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I2014 WorkList ID : 96004 Date : 2/27/2017 3:56:20 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2014-01	Corrosivity	Cool 4 deg C	ARAM01	I62	PRESS-WASTE-WATER-CAK	02/27/2017	9045C

Date/Time 02/27/17 3:56 PM
 Received by: JD
 Relinquished by: JD

Date/Time 02/27/17 5:30 PM
 Received by: JD
 Relinquished by: JD