


ISM02.4

LB 86066
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44515



A= 0.0000e+000

B= 2.2484e-004

C= 3.3828e-002

Rho= 0.9999332

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.036	0.036	9	0.000	9					-
0.2	0.200	0.230	0.030	872	0.0 %	872					15
2.5	2.500	2.425	-0.075	10634	0.0 %	10634					-3
5.0	5.000	5.003	0.003	22099	0.0 %	22099					0
7.5	7.500	7.464	-0.036	33049	0.0 %	33049					0
10.0	10.000	10.042	0.042	44515	0.0 %	44515					0

mB
2/28/17

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	9	-		0 ISM02.4	PPB	2/28/2017 11:12	Std
	0.2 S0.2	872	-		0.2 ISM02.4	PPB	2/28/2017 11:14	Std
	2.5 S2.5	10634	-		2.5 ISM02.4	PPB	2/28/2017 11:19	Std
	5 S5	22099	-		5 ISM02.4	PPB	2/28/2017 11:21	Std
	7.5 S7.5	33049	-		7.5 ISM02.4	PPB	2/28/2017 11:23	Std
	10 S10	44515	-		10 ISM02.4	PPB	2/28/2017 11:25	Std
ICV	ICV	17668	4.0063 -		ISM02.4	PPB	2/28/2017 11:29	SMPL
ICB	ICB	-72	0.0176 -		ISM02.4	PPB	2/28/2017 11:31	SMPL
CCV001	CCV001	22149	5.0138 -		ISM02.4	PPB	2/28/2017 11:34	SMPL
CCB001	CCB001	2	0.0343 -		ISM02.4	PPB	2/28/2017 11:38	SMPL
PB97209BL	PBW003	5	0.035 -		ISM02.4	PPB	2/28/2017 11:41	SMPL
I1984-01	MBC863	193	0.0772 -		ISM02.4	PPB	2/28/2017 11:43	SMPL
I1984-02	MBC863D	66	0.0487 -		ISM02.4	PPB	2/28/2017 11:45	SMPL
I1984-03	MBC863S	4536	1.0537 -		ISM02.4	PPB	2/28/2017 11:50	SMPL
I1984-04	MBC864	43	0.0435 -		ISM02.4	PPB	2/28/2017 11:52	SMPL
I1984-05	MBC865	57	0.0466 -		ISM02.4	PPB	2/28/2017 11:55	SMPL
CCV002	CCV002	22251	5.0367 -		ISM02.4	PPB	2/28/2017 11:57	SMPL
CCB002	CCB002	-61	0.0201 -		ISM02.4	PPB	2/28/2017 11:59	SMPL