

7470A


L986071

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36145



A= 0.0000e+000

B= 2.7383e-004

C= 5.4168e-003

Rho= 0.9997926

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.004	0.004	-6	0.000	-6				
0.2	0.200	0.192	-0.008	681	0.0 %	681				
2.5	2.500	2.480	-0.020	9036	0.0 %	9036				
5.0	5.000	4.969	-0.031	18128	0.0 %	18128				
7.5	7.500	7.652	0.152	27924	0.0 %	27924				
10.0	10.000	9.903	-0.097	36145	0.0 %	36145				

LB86071 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 28 Feb 2017 12:40:42				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	S0	-6	-	0.0000	7470A	PPB	28 Feb 2017 12:42:43	S Std
0.2 - 1	S0.2	681	-	0.2000	7470A	PPB	28 Feb 2017 12:44:51	S Std
2.5 - 1	S2.5	9036	-	2.5000	7470A	PPB	28 Feb 2017 12:47:00	S Std
5.0 - 1	S5	18128	-	5.0000	7470A	PPB	28 Feb 2017 12:49:08	S Std
7.5 - 1	S7.5	27924	-	7.5000	7470A	PPB	28 Feb 2017 12:51:16	S Std
10.0 - 1	S10	36145	-	10.0000	7470A	PPB	28 Feb 2017 12:56:03	S Std
ICV51 - 1	ICV51	14395	3.9473	-	7470A	PPB	28 Feb 2017 12:58:37	U SMPL
ICB51 - 1	ICB51	-122	-0.0280	-	7470A	PPB	28 Feb 2017 13:00:43	U SMPL
CCV03 - 1	CCV03	17824	4.8862	-	7470A	PPB	28 Feb 2017 13:02:50	U SMPL
CCB03 - 1	CCB03	-125	-0.0288	-	7470A	PPB	28 Feb 2017 13:04:57	U SMPL
CRA - 1	CRA	639	0.1804	-	7470A	PPB	28 Feb 2017 13:07:03	U SMPL
HighStd - 1	HighStd	35715	9.7854	-	7470A	PPB	28 Feb 2017 13:09:10	U SMPL
ChkStd - 1	ChkStd	24783	6.7918	-	7470A	PPB	28 Feb 2017 13:11:17	U SMPL
PB97216BL - 1	PBW	-156	-0.0373	-	7470A	PPB	28 Feb 2017 13:13:25	U SMPL
PB97216BS - 1	LCSW	13894	3.8101	-	7470A	PPB	28 Feb 2017 13:15:33	U SMPL
I1996-08 - 1	E2-FB-2-24	-95	-0.0206	-	7470A	PPB	28 Feb 2017 13:17:40	U SMPL
I2011-01 - 1	FRAC-TANK-A155	1084	0.3023	-	7470A	PPB	28 Feb 2017 13:19:48	U SMPL
I2011-01DUP - 1	FRAC-TANK-A155DUP	937	0.2620	-	7470A	PPB	28 Feb 2017 13:21:57	U SMPL
I2011-01MS - 1	FRAC-TANK-A155MS	14743	4.0425	-	7470A	PPB	28 Feb 2017 13:24:05	U SMPL
I2011-01MSD - 1	FRAC-TANK-A155MSD	13898	3.8112	-	7470A	PPB	28 Feb 2017 13:26:12	U SMPL
CCV04 - 1	CCV04	17839	4.8903	-	7470A	PPB	28 Feb 2017 13:28:19	U SMPL
CCB04 - 1	CCB04	-83	-0.0173	-	7470A	PPB	28 Feb 2017 13:30:26	U SMPL
PB97217BL - 1	PBW	-23	-0.0009	-	7470A	PPB	28 Feb 2017 13:32:33	U SMPL
PB97217BS - 1	LCSW	12831	3.5190	-	7470A	PPB	28 Feb 2017 13:34:40	U SMPL
PB97217TB - 1	PB97181TB	-83	-0.0173	-	7470A	PPB	28 Feb 2017 13:36:46	U SMPL
I1998-03 - 1	ROLL-OFF-15-50-COMP	-31	-0.0031	-	7470A	PPB	28 Feb 2017 13:38:53	U SMPL
I1998-06 - 1	ROLL-OFF-15-41-COMP	253	0.0747	-	7470A	PPB	28 Feb 2017 13:41:01	U SMPL
I1998-09 - 1	ROLL-OFF-20-41-COMP	-13	0.0019	-	7470A	PPB	28 Feb 2017 13:43:08	U SMPL
I1998-12 - 1	ROLL-OFF-20-41-COMP	-5	0.0040	-	7470A	PPB	28 Feb 2017 13:45:16	U SMPL
I2000-03 - 1	SP-NORTH(0-2)	57	0.0210	-	7470A	PPB	28 Feb 2017 13:47:24	U SMPL
I2000-06 - 1	SP-SOUTH(0-2)	1	0.0057	-	7470A	PPB	28 Feb 2017 13:49:33	U SMPL
I2000-06DUP - 1	SP-SOUTH(0-2)DUP	2	0.0060	-	7470A	PPB	28 Feb 2017 13:51:40	U SMPL
CCV05 - 1	CCV05	17766	4.8703	-	7470A	PPB	28 Feb 2017 13:53:47	U SMPL
CCB05 - 1	CCB05	-92	-0.0198	-	7470A	PPB	28 Feb 2017 13:55:54	U SMPL
I2000-06MS - 1	SP-SOUTH(0-2)MS	12768	3.5017	-	7470A	PPB	28 Feb 2017 13:58:00	U SMPL
I2000-06MSD - 1	SP-SOUTH(0-2)MSD	12901	3.5381	-	7470A	PPB	28 Feb 2017 14:00:07	U SMPL
I2011-01LX5 - 1		119	0.0380	-	7470A	PPB	28 Feb 2017 14:02:14	U SMPL
I2011-01A - 1		14530	3.9842	-	7470A	PPB	28 Feb 2017 14:04:21	U SMPL
I2000-06LX5 - 1		-80	-0.0165	-	7470A	PPB	28 Feb 2017 14:06:28	U SMPL
I2000-06A - 1		14345	3.9336	-	7470A	PPB	28 Feb 2017 14:08:36	U SMPL
CCV06 - 1	CCV06	17684	4.8479	-	7470A	PPB	28 Feb 2017 14:10:44	U SMPL
CCB06 - 1	CCB06	-117	-0.0266	-	7470A	PPB	28 Feb 2017 14:12:52	U SMPL