

7471A


LB86091

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46440



A= 0.0000e+000

B= 2.1540e-004

C= -7.4275e-003

Rho= 0.9999593

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.011	0.011	86	0.000	86				
0.2	0.200	0.185	-0.015	892	0.0 %	892				
2.5	2.500	2.536	0.036	11807	0.0 %	11807				
5.0	5.000	4.938	-0.062	22960	0.0 %	22960				
7.5	7.500	7.534	0.034	35012	0.0 %	35012				
10.0	10.000	9.996	-0.004	46440	0.0 %	46440				

LB86091 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 01 Mar 2017 07:52:59

Sample ID	Extended ID	µ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	86	-	0.0000	7471A	PPB	01 Mar 2017 07:58:42	S	Std
0.2 - 1	50.2	892	-	0.2000	7471A	PPB	01 Mar 2017 08:00:59	S	Std
2.5 - 1	52.5	11807	-	2.5000	7471A	PPB	01 Mar 2017 08:03:16	S	Std
5.0 - 1	55	22960	-	5.0000	7471A	PPB	01 Mar 2017 08:05:34	S	Std
7.5 - 1	57.5	35012	-	7.5000	7471A	PPB	01 Mar 2017 08:07:51	S	Std
10.0 - 1	510	46440	-	10.0000	7471A	PPB	01 Mar 2017 08:10:07	S	Std
ICV52 - 1	ICV52	17483	3.7585	-	7471A	PPB	01 Mar 2017 08:15:28	U	SMPL
ICB52 - 1	ICB52	22	-0.0027	-	7471A	PPB	01 Mar 2017 08:17:43	U	SMPL
CCV09 - 1	CCV09	21790	4.6862	-	7471A	PPB	01 Mar 2017 08:19:59	U	SMPL
CCB09 - 1	CCB09	14	-0.0044	-	7471A	PPB	01 Mar 2017 08:22:15	U	SMPL
CRA - 1	CRA	904	0.1873	-	7471A	PPB	01 Mar 2017 08:24:31	U	SMPL
HighStd - 1	HighStd	46550	10.0196	-	7471A	PPB	01 Mar 2017 08:26:47	U	SMPL
ChkStd - 1	ChkStd	31870	6.8575	-	7471A	PPB	01 Mar 2017 08:29:03	U	SMPL
PB97223BL - 1	PBS	-20	-0.0117	-	7471A	PPB	01 Mar 2017 08:31:20	U	SMPL
PB97223BS - 1	LCSS	17020	3.6587	-	7471A	PPB	01 Mar 2017 08:33:37	U	SMPL
I1996-01 - 1	E2-AA5-SWW	96251	20.7253	-	7471A	PPB	01 Mar 2017 08:35:54	U	SMPL
I1996-02 - 1	E2-AA6-BASE	4861	1.0396	-	7471A	PPB	01 Mar 2017 08:38:11	U	SMPL
I1996-02DUP - 1	E2-AA6-BASEDUP	4950	1.0588	-	7471A	PPB	01 Mar 2017 08:40:30	U	SMPL
I1996-03 - 1	I1996-02MS	20182	4.3398	-	7471A	PPB	01 Mar 2017 08:42:48	U	SMPL
I1996-04 - 1	I1996-02MSD	20391	4.3849	-	7471A	PPB	01 Mar 2017 08:45:04	U	SMPL
CCV10 - 1	CCV10	21932	4.7168	-	7471A	PPB	01 Mar 2017 08:47:20	U	SMPL
CCB10 - 1	CCB10	24	-0.0023	-	7471A	PPB	01 Mar 2017 08:49:36	U	SMPL
I1996-05 - 1	E2-AA6-SWN	31964	6.8777	-	7471A	PPB	01 Mar 2017 08:51:52	U	SMPL
I1996-06 - 1	E2-AA6-SWW	7593	1.6281	-	7471A	PPB	01 Mar 2017 08:54:08	U	SMPL
I1996-07 - 1	E2-AA5D-SWW	146972	31.6508	-	7471A	PPB	01 Mar 2017 08:56:24	U	SMPL
I1997-01 - 1	E2-W5-BASE	7406	1.5878	-	7471A	PPB	01 Mar 2017 08:58:41	U	SMPL
I1997-02 - 1	E2-W4-WSW	26821	5.7699	-	7471A	PPB	01 Mar 2017 09:01:00	U	SMPL
I1997-03 - 1	E2-W6-BASE	337472	HIGH	-	7471A	PPB	01 Mar 2017 09:03:16	X	
I1997-04 - 1	E2-W7-BASE	63728	13.7198	-	7471A	PPB	01 Mar 2017 09:05:07	U	SMPL
I1997-05 - 1	E2-BB6-BASE	18800	4.0421	-	7471A	PPB	01 Mar 2017 09:07:26	U	SMPL
I1997-06 - 1	E2-BB6-WSW	748	0.1537	-	7471A	PPB	01 Mar 2017 09:09:45	U	SMPL
I1997-07 - 1	E2-B-1E-NSW-1	315103	HIGH	-	7471A	PPB	01 Mar 2017 09:12:01	X	
CCV11 - 1	CCV11	21779	4.6838	-	7471A	PPB	01 Mar 2017 09:13:52	U	SMPL
CCB11 - 1	CCB11	6	-0.0061	-	7471A	PPB	01 Mar 2017 09:16:10	U	SMPL
I1997-08 - 1	E2-B1E-ESW-2	138928	29.9181	-	7471A	PPB	01 Mar 2017 09:18:27	U	SMPL
I1997-09 - 1	E2-B1E-WSW-2	311899	67.1765	-	7471A	PPB	01 Mar 2017 09:20:43	U	SMPL
I1997-10 - 1	E2-B1E-ESW-1	91230	19.6438	-	7471A	PPB	01 Mar 2017 09:23:01	U	SMPL
I1997-11 - 1	E2-B1E-WSW-1	10466	2.2470	-	7471A	PPB	01 Mar 2017 09:25:20	U	SMPL
I1997-12 - 1	E2-B1E-SSW-1	30443	6.5501	-	7471A	PPB	01 Mar 2017 09:27:38	U	SMPL
I1996-02LX5 - 1		963	0.2000	-	7471A	PPB	01 Mar 2017 09:29:55	U	SMPL
I1996-02A - 1		21922	4.7146	-	7471A	PPB	01 Mar 2017 09:32:13	U	SMPL
CCV12 - 1	CCV12	22149	4.7635	-	7471A	PPB	01 Mar 2017 09:34:30	U	SMPL
CCB12 - 1	CCB12	30	-0.0010	-	7471A	PPB	01 Mar 2017 09:36:48	U	SMPL
I1996-01DLX5 - 1		19291	4.1479	-	7471A	PPB	01 Mar 2017 09:39:04	U	SMPL
I1996-07DLX5 - 1		30833	6.6341	-	7471A	PPB	01 Mar 2017 09:41:20	U	SMPL
I1997-03DLX10 - 1		47721	10.2718	-	7471A	PPB	01 Mar 2017 09:43:37	U	SMPL
I1997-04DLX5 - 1		13157	2.8266	-	7471A	PPB	01 Mar 2017 09:45:53	U	SMPL
I1997-07DLX10 - 1		41157	8.8579	-	7471A	PPB	01 Mar 2017 09:48:09	U	SMPL
I1997-08DLX5 - 1		28296	6.0876	-	7471A	PPB	01 Mar 2017 09:50:26	U	SMPL
I1997-09DLX10 - 1		34432	7.4093	-	7471A	PPB	01 Mar 2017 09:52:42	U	SMPL
I1997-10DLX5 - 1		18654	4.0107	-	7471A	PPB	01 Mar 2017 09:54:59	U	SMPL
CCV13 - 1	CCV13	22680	4.8779	-	7471A	PPB	01 Mar 2017 09:57:16	U	SMPL
CCB13 - 1	CCB13	32	-0.0005	-	7471A	PPB	01 Mar 2017 09:59:33	U	SMPL
I1997-03DL2X50 - 1		9306	1.9971	-	7471A	PPB	01 Mar 2017 10:01:50	U	SMPL
I1973-18DL2X100 - 1		26918	5.7908	-	7471A	PPB	01 Mar 2017 10:04:08	U	SMPL
CCV14 - 1	CCV14	23042	4.9559	-	7471A	PPB	01 Mar 2017 10:06:25	U	SMPL
CCB14 - 1	CCB14	45	0.0023	-	7471A	PPB	01 Mar 2017 10:08:41	U	SMPL