



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86095

Temp.: 22.3 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/01/2017	11:50
2	CAL2	1	Water			7.00	03/01/2017	11:51
3	CAL3	1	Water			10.03	03/01/2017	11:52
4	ICV	1	Water			6.99	03/01/2017	11:53
5	CCV1	1	Water			2.01	03/01/2017	11:55
6	I2040-01	1	Water			4.64	03/01/2017	11:56
7	I2040-01DUP	1	Water			4.65	03/01/2017	11:57
8	CCV2	1	Water			12.02	03/01/2017	11:58

1386095

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i2040

WorkList ID : 96073

Date : 3/1/2017 11:37:46 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2040-01	pH	Cool 4 deg C	STER14	J11	POOLWATER	02/24/2017	SM 4500-PH B

Date/Time 03.01.17 11:37 AM
Received by: JP
Relinquished by: [Signature]

Date/Time 03.01.17 [Signature]
Received by: [Signature]
Relinquished by: SP