

Chemtech Consulting Group

Analytical Review Report

Date Printed : 3/2/17

Analyst : AK

Data File : LB86106.csv

Approved By :

Approved Date :

Worksheet # :


m300.0 - Inorganic Anions - 16

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1	
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Bromide												
STD1	STD1				W	2/14/17						
Bromide		PASS	0.000		0.000						mg/L	
STD2	STD2				W	2/14/17						
Bromide		PASS	0.895		0.895						mg/L	
STD3	STD3				W	2/14/17						
Bromide		PASS	2.264		2.26						mg/L	
STD4	STD4				W	2/14/17						
Bromide		PASS	4.074		4.07						mg/L	
STD5	STD5				W	2/14/17						
Bromide		PASS	9.757		9.76						mg/L	
STD6	STD6				W	2/14/17						
Bromide		PASS	19.517		19.5						mg/L	
STD7	STD7				W	2/14/17						
Bromide		PASS	24.577		24.6						mg/L	
STD8	STD8				W	2/14/17						
Bromide		PASS	50.415		50.4						mg/L	
ICV1	ICV1				W	2/14/17						
Bromide		PASS	25.120		25.10	100.0	90	110			mg/L	
ICB1	ICB1				W	2/14/17						
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L	
CCV1	CCV1				W	3/1/17						
Bromide		PASS	25.442		25.40	102.0	90	110			mg/L	
CCB1	CCB1				W	3/1/17						
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L	
LB86106BLW	LB86106BLW				W	3/1/17						
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L	
LB86106BSW	LB86106BSW				W	3/1/17						
Bromide		PASS	24.137		24.10	96.0	90	110			mg/L	
I2040-01	POOLWATER		1		W	3/1/17						
Bromide		PASS	0.000		0.000						mg/L	
I2040-01D	POOLWATERD		1		W	3/1/17						
Bromide		PASS	0.000		0.00				0	20	mg/L	
I2040-01S	POOLWATERS		1		W	3/1/17						
Bromide		PASS	23.864		23.90	96.0	80	120			mg/L	
CCV2	CCV2				W	3/1/17						
Bromide		PASS	24.734		24.70	99.0	90	110			mg/L	
CCB2	CCB2				W	3/1/17						
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 3/2/17
 Analyst : AK
 Data File : LB86106.csv


Approved By : 
 Approved Date :
 Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Chloride												
STD1	STD1				W	2/14/17						
Chloride		PASS	0.000		0.000						mg/L	
STD2	STD2				W	2/14/17						
Chloride		PASS	0.366		0.366						mg/L	
STD3	STD3				W	2/14/17						
Chloride		PASS	0.735		0.735						mg/L	
STD4	STD4				W	2/14/17						
Chloride		PASS	1.270		1.27						mg/L	
STD5	STD5				W	2/14/17						
Chloride		PASS	2.955		2.96						mg/L	
STD6	STD6				W	2/14/17						
Chloride		PASS	5.688		5.69						mg/L	
STD7	STD7				W	2/14/17						
Chloride		PASS	7.176		7.18						mg/L	
STD8	STD8				W	2/14/17						
Chloride		PASS	15.260		15.3						mg/L	
ICV1	ICV1				W	2/14/17						
Chloride		PASS	7.104		7.10	95.0	90	110			mg/L	
ICB1	ICB1				W	2/14/17						
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L	
CCV1	CCV1				W	3/1/17						
Chloride		PASS	7.050		7.05	94.0	90	110			mg/L	
CCB1	CCB1				W	3/1/17						
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L	
LB86106BLW	LB86106BLW				W	3/1/17						
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L	
LB86106BSW	LB86106BSW				W	3/1/17						
Chloride		PASS	6.910		6.91	92.0	90	110			mg/L	
I2040-01	POOLWATER		1		W	3/1/17						
Chloride		PASS	0.000		0.000						mg/L	
I2040-01D	POOLWATERD		1		W	3/1/17						
Chloride		PASS	0.000		0.00				0	20	mg/L	
I2040-01S	POOLWATERS		1		W	3/1/17						
Chloride		PASS	6.933		6.93	92.0	80	120			mg/L	
CCV2	CCV2				W	3/1/17						
Chloride		PASS	7.175		7.18	96.0	90	110			mg/L	
CCB2	CCB2				W	3/1/17						
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 3/2/17
 Analyst : AK
 Data File : LB86106.csv

Approved By : 
 Approved Date : 
 Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1	
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Fluoride												
STD1	STD1				W	2/14/17						
Fluoride		PASS	0.000		0.000						mg/L	
STD2	STD2				W	2/14/17						
Fluoride		PASS	0.224		0.224						mg/L	
STD3	STD3				W	2/14/17						
Fluoride		PASS	0.450		0.450						mg/L	
STD4	STD4				W	2/14/17						
Fluoride		PASS	0.762		0.762						mg/L	
STD5	STD5				W	2/14/17						
Fluoride		PASS	1.927		1.93						mg/L	
STD6	STD6				W	2/14/17						
Fluoride		PASS	3.924		3.92						mg/L	
STD7	STD7				W	2/14/17						
Fluoride		PASS	4.943		4.94						mg/L	
STD8	STD8				W	2/14/17						
Fluoride		PASS	10.069		10.1						mg/L	
ICV1	ICV1				W	2/14/17						
Fluoride		PASS	5.215		5.22	104.0	90	110			mg/L	
ICB1	ICB1				W	2/14/17						
Fluoride		PASS	0.000		0.000			+/-0.0430			mg/L	
CCV1	CCV1				W	3/1/17						
Fluoride		PASS	5.227		5.23	105.0	90	110			mg/L	
CCB1	CCB1				W	3/1/17						
Fluoride		PASS	0.000		0.000			+/-0.0430			mg/L	
LB86106BLW	LB86106BLW				W	3/1/17						
Fluoride		PASS	0.000		0.000			+/-0.0430			mg/L	
LB86106BSW	LB86106BSW				W	3/1/17						
Fluoride		PASS	5.163		5.16	103.0	90	110			mg/L	
I2040-01	POOLWATER	1			W	3/1/17						
Fluoride		PASS	0.000		0.000						mg/L	
I2040-01D	POOLWATERD	1			W	3/1/17						
Fluoride		PASS	0.000		0.00				0	20	mg/L	
I2040-01S	POOLWATERS	1			W	3/1/17						
Fluoride		PASS	5.936		5.94	119.0	80	120			mg/L	
CCV2	CCV2				W	3/1/17						
Fluoride		PASS	5.407		5.41	108.0	90	110			mg/L	
CCB2	CCB2				W	3/1/17						
Fluoride		PASS	0.000		0.000			+/-0.0430			mg/L	