

Run Number: LB86113

BalanceID: WC SC-3

DATE & TIME IN: 03/02/2017 10:10

DATE & TIME OUT: 03/02/2017 14:25

OVEN TEMP: 104 °C

OVEN TEMP: 105 °C

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Weight (g)	Result mg/L
1	LB86113BL	LB86113BL	1.4823	1.4823	100	1.4823	1.4823	1.4823	0	0
2	LB86113BS	LB86113BS	1.4025	1.4025	100	1.4573	1.4573	1.4573	0.0548	548
3	I2029-01	BAYPARK-122816	1.4044	1.4044	400	1.4055	1.4055	1.4055	0.0011	2.8
4	I2039-01	DRGFB-WQ-065-BG-A	1.426	1.426	400	1.4393	1.4393	1.4393	0.0133	33.2
5	I2039-02	DRGFB-WQ-066-AD-A	1.4083	1.4083	200	1.4147	1.4147	1.4147	0.0064	32
6	I2039-02DUP	DRGFB-WQ-066-AD-ADUP	1.4238	1.4238	200	1.4305	1.4305	1.4305	0.0067	33.5
7	I2042-01	COMP	1.4144	1.4144	200	1.4459	1.4459	1.4459	0.0315	157.5

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

$$\text{Weight (g)} = C - B$$

$$\text{Result mg/L} = \frac{D}{A} * 1000 * 1000$$

1386113

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS I2042

WorkList ID : 96092

Date : 3/2/2017 8:58:24 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2029-01	TSS	Cool 4 deg C	BSRJ01	I51	BAYPARK-122816	02/27/2017	SM2540 D
	Water	I2039-01-1	TSS	Cool 4 deg C	JAYC01	J12	DRGFB-WQ-065-BG-A	02/28/2017	SM2540 D
	Water	I2039-02-1	TSS	Cool 4 deg C	JAYC01	J12	DRGFB-WQ-066-AD-A	02/28/2017	SM2540 D
	Water	I2042-01-1	TSS	Cool 4 deg C	ARAM01	J12	COMP	03/01/2017	SM2540 D

Date/Time 03/02/17 8:58 AM
 Received by: JD
 Relinquished by: CP

Date/Time 03/02/17 11:30 AM
 Received by: CP
 Relinquished by: JD