



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB86114
BalanceID: WC SC-3
Analysis Date: 03/02/2017

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 03/02/2017
Extraction IN Time: 09:10
Extraction OUT Time: 09:40

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight	Weight After Drying	Final Weight After Drying	Change Weight (g)	Result in ppm
1	LB86114BL	LB86114BL	WATER	2	1000	100	3.1542	3.1542	3.1542	3.1542	0	0
2	LB86114BS	LB86114BS	WATER	2	1000	100	3.8501	3.8501	3.8692	3.8692	0.0191	19.1
3	LB86114BSD	LB86114BSD	WATER	2	1000	100	3.4879	3.4879	3.5073	3.5073	0.0194	19.4
4	I2042-02	GRAB	WATER	2	1000	100	3.131	3.131	3.1597	3.1597	0.0287	28.7



QC Batch# LB86114 Test: TPH Analysis Date: 03/02/2017

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2203
pH Paper 0-14	W2105
Sodium Sulfate	
1:1 HCL	WP50630
Silica Gel	W2061 (3g)
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP50631
LCSWD	10.00 ML	WP50631
MS/MSD		

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance: 0.0018 In OVEN TEMP : 70 Dessicator Time In : 11:00
1.0000 gram Balance: 1.0005 In Time: 10:30
Bal Check Time: 09:30

After Analysis

0.0020 gram Balance: 0.0021 Out OVEN TEMP: 71 Dessicator Time Out: 13:00
1.0000 gram Balance: 1.0003 Out Time: 12:10
Bal Check Time: 13:30

JB 86114

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH I2042

WorkList ID : 96093

Date : 3/2/2017 8:59:11 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2042-02	TPH	Conc H2SO4 to	ARAM01	J12	GRAB	03/01/2017	1664A

Date/Time 03/02/17 8:59 AM
Received by: JB
Relinquished by: CP

Date/Time 03/02/17 11:00 AM
Received by: CP
Relinquished by: JB