


7471A

LB 86116
INSTRUMENT ID: CV2

Linear

 μ Abs.:

4284E



A= 0.0000e+000

B= 2.3271e-004

C= -2.2437e-002

Rho= 0.9999558

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.007	-0.007	68	0.000	68				
0.2	0.200	0.170	-0.030	829	0.0 %	829				
2.5	2.500	2.504	0.004	10858	0.0 %	10858				
5.0	5.000	5.048	0.048	21791	0.0 %	21791				
7.5	7.500	7.535	0.035	32478	0.0 %	32478				
10.0	10.000	9.948	-0.052	42846	0.0 %	42846				

LB86116 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 02 Mar 2017 10:19:14

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	68	-	0.0000	7471A	PPB	02 Mar 2017 10:27:30	S	Std
0.2 - 1	50.2	829	-	0.2000	7471A	PPB	02 Mar 2017 10:29:48	S	Std
2.5 - 1	52.5	10858	-	2.5000	7471A	PPB	02 Mar 2017 10:32:05	S	Std
5.0 - 1	55	21791	-	5.0000	7471A	PPB	02 Mar 2017 10:34:23	S	Std
7.5 - 1	57.5	32478	-	7.5000	7471A	PPB	02 Mar 2017 10:36:40	S	Std
10.0 - 1	57.5	42846	-	10.0000	7471A	PPB	02 Mar 2017 10:41:30	S	Std
ICV54 - 1	ICV54	17440	4.0359	-	7471A	PPB	02 Mar 2017 10:44:14	U	SMPL
ICB54 - 1	ICB54	-46	-0.0331	-	7471A	PPB	02 Mar 2017 10:46:29	U	SMPL
CCV18 - 1	CCV18	21261	4.9251	-	7471A	PPB	02 Mar 2017 10:48:45	U	SMPL
CCB18 - 1	CCB18	-60	-0.0364	-	7471A	PPB	02 Mar 2017 10:51:00	U	SMPL
CRA - 1	CRA	798	0.1633	-	7471A	PPB	02 Mar 2017 10:53:16	U	SMPL
HighStd - 1	HighStd	43510	10.1026	-	7471A	PPB	02 Mar 2017 10:55:31	U	SMPL
ChkStd - 1	ChkStd	31991	7.4220	-	7471A	PPB	02 Mar 2017 10:57:48	U	SMPL
PB97280BL - 1	PBS	-58	-0.0359	-	7471A	PPB	02 Mar 2017 11:00:04	U	SMPL
PB97280BS - 1	LCSS	17039	3.9426	-	7471A	PPB	02 Mar 2017 11:02:21	U	SMPL
I2015-02 - 1	37A-9-11	1730	0.3801	-	7471A	PPB	02 Mar 2017 11:04:38	U	SMPL
I2015-06 - 1	38A-7-9	1212	0.2596	-	7471A	PPB	02 Mar 2017 11:06:56	U	SMPL
I2033-01 - 1	FK-GF-03-002-A	1746	0.3839	-	7471A	PPB	02 Mar 2017 11:09:13	U	SMPL
I2033-03 - 1	FK-GF-03-002-B	1825	0.4022	-	7471A	PPB	02 Mar 2017 11:11:31	U	SMPL
I2034-01 - 1	FK-BPM-01-DD9-DD10-DD11-DD12	349	0.0588	-	7471A	PPB	02 Mar 2017 11:13:47	U	SMPL
CCV19 - 1	CCV19	21087	4.8846	-	7471A	PPB	02 Mar 2017 11:16:03	U	SMPL
CCB19 - 1	CCB19	-28	-0.0290	-	7471A	PPB	02 Mar 2017 11:18:19	U	SMPL
I2035-01 - 1	FK-BPM-01-EE9-EE10-EE11-EE12	381	0.0662	-	7471A	PPB	02 Mar 2017 11:20:35	U	SMPL
I2036-01 - 1	FK-GF-03-001-A	1767	0.3888	-	7471A	PPB	02 Mar 2017 11:22:51	U	SMPL
I2036-03 - 1	FK-GF-03-001-B	1528	0.3331	-	7471A	PPB	02 Mar 2017 11:25:07	U	SMPL
I2036-05 - 1	FK-GF-03-001-C	1343	0.2901	-	7471A	PPB	02 Mar 2017 11:27:23	U	SMPL
I2037-01 - 1	FK-BPM-01-BB5-BB6-BB7-BB8	411	0.0732	-	7471A	PPB	02 Mar 2017 11:29:40	U	SMPL
I2037-01DUP - 1	FK-BPM-01-BB5-BB6-BB7-BB8DUP	354	0.0599	-	7471A	PPB	02 Mar 2017 11:31:57	U	SMPL
I2037-01MS - 1	FK-BPM-01-BB5-BB6-BB7-BB8MS	16455	3.8067	-	7471A	PPB	02 Mar 2017 11:34:14	U	SMPL
I2037-01MSD - 1	FK-BPM-01-BB5-BB6-BB7-BB8MSD	16035	3.7090	-	7471A	PPB	02 Mar 2017 11:36:31	U	SMPL
I2037-01LX5 - 1		132	0.0083	-	7471A	PPB	02 Mar 2017 11:40:39	U	SMPL
I2037-01A - 1		17650	4.0848	-	7471A	PPB	02 Mar 2017 11:42:54	U	SMPL
CCV20 - 1	CCV20	21382	4.9533	-	7471A	PPB	02 Mar 2017 11:45:10	U	SMPL
CCB20 - 1	CCB20	-35	-0.0306	-	7471A	PPB	02 Mar 2017 11:47:26	U	SMPL