


245.1

LB86118
INSTRUMENT ID: CV2

Linear

 μ Abs.:

33276



A= 0.0000e+000

B= 2.9874e-004

C= -3.5625e-002

Rho= 0.9998511

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.034	-0.034	6	0.000	6				
0.2	0.200	0.163	-0.037	664	0.0 %	664				
2.5	2.500	2.557	0.057	8678	0.0 %	8678				
5.0	5.000	5.012	0.012	16895	0.0 %	16895				
7.5	7.500	7.597	0.097	25550	0.0 %	25550				
10.0	10.000	9.905	-0.095	33276	0.0 %	33276				

LB86118 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 02 Mar 2017 12:20:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	6	-	0.0000	245.1	PPB	02 Mar 2017 12:25:43	S	Std
0.2 - 1	50.2	664	-	0.2000	245.1	PPB	02 Mar 2017 12:27:51	S	Std
2.5 - 1	52.5	8678	-	2.5000	245.1	PPB	02 Mar 2017 12:30:00	S	Std
5.0 - 1	55	16895	-	5.0000	245.1	PPB	02 Mar 2017 12:32:08	S	Std
7.5 - 1	57.5	25550	-	7.5000	245.1	PPB	02 Mar 2017 12:34:16	S	Std
10.0 - 1	510	33276	-	10.0000	245.1	PPB	02 Mar 2017 12:36:23	S	Std
ICV55 - 1	ICV55	13778	4.0805	-245.1		PPB	02 Mar 2017 12:39:25	U	SMPL
ICB55 - 1	ICB55	-53	-0.0515	-245.1		PPB	02 Mar 2017 12:41:31	U	SMPL
CCV21 - 1	CCV21	16649	4.9381	-245.1		PPB	02 Mar 2017 12:43:37	U	SMPL
CCB21 - 1	CCB21	-61	-0.0538	-245.1		PPB	02 Mar 2017 12:45:44	U	SMPL
CRA - 1	CRA	715	0.1780	-245.1		PPB	02 Mar 2017 12:47:50	U	SMPL
HighStd - 1	HighStd	32527	9.6816	-245.1		PPB	02 Mar 2017 12:49:57	U	SMPL
ChkStd - 1	ChkStd	23412	6.9585	-245.1		PPB	02 Mar 2017 12:52:04	U	SMPL
PB97273BL - 1	PBW	-72	-0.0571	-245.1		PPB	02 Mar 2017 12:54:12	U	SMPL
PB97273BS - 1	LCSW	13643	4.0401	-245.1		PPB	02 Mar 2017 12:56:19	U	SMPL
I2009-01 - 1	022417-PRRT-F-D-M	26	-0.0279	-245.1		PPB	02 Mar 2017 12:58:27	U	SMPL
I2009-02 - 1	022417-PRRT-F-U-M	68	-0.0153	-245.1		PPB	02 Mar 2017 13:00:35	U	SMPL
I2009-02DUP - 1	022417-PRRT-F-U-MDUP	60	-0.0177	-245.1		PPB	02 Mar 2017 13:02:43	U	SMPL
I2009-02MS - 1	022417-PRRT-F-U-MMS	14301	4.2367	-245.1		PPB	02 Mar 2017 13:07:24	U	SMPL
I2009-02MSD - 1	022417-PRRT-F-U-MMSD	14241	4.2188	-245.1		PPB	02 Mar 2017 13:09:30	U	SMPL
CCV22 - 1	CCV22	16721	4.9597	-245.1		PPB	02 Mar 2017 13:11:37	U	SMPL
CCB22 - 1	CCB22	-55	-0.0521	-245.1		PPB	02 Mar 2017 13:13:44	U	SMPL
I2042-01 - 1	COMP	248	0.0385	-245.1		PPB	02 Mar 2017 13:15:51	U	SMPL
I2009-02LX5 - 1		3	-0.0347	-245.1		PPB	02 Mar 2017 13:17:58	U	SMPL
I2009-02A - 1		13796	4.0858	-245.1		PPB	02 Mar 2017 13:20:05	U	SMPL
CCV23 - 1	CCV23	16824	4.9904	-245.1		PPB	02 Mar 2017 13:22:12	U	SMPL
CCB23 - 1	CCB23	-80	-0.0595	-245.1		PPB	02 Mar 2017 13:24:20	U	SMPL