

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB86123

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP50397
calibration std. phosphate 1 ppm	WP53287
calibration std. phosphate 0.5 ppm	WP53286
calibration std. phosphate 0.3 ppm	WP53285
calibration std. 0 ppm	WP53282
phosphate ICV std.	WP53289
phosphate CCV std.	WP53288
calibration std. phosphate 0.1 ppm	WP53284
calibration std. phosphate 0.05 ppm	WP53283
Combined reagent	WP53291
Phenolphthalein indicator	WP52748
Sodium hydroxide, 1N	WP52749
Ascorbic Acid	WP53290

Intercept: -0.0001Slope: 0.6543Regression: 0.999934

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume	Final Volume	Reading	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	03/02/2017	14:35
2	CAL2	0.05	1	50	50	0.033	03/02/2017	14:45
3	CAL3	0.10	1	50	50	0.062	03/02/2017	14:50
4	CAL4	0.30	1	50	50	0.196	03/02/2017	14:50
5	CAL5	0.50	1	50	50	0.332	03/02/2017	14:51
6	CAL6	1.00	1	50	50	0.652	03/02/2017	14:51



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB86123

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol/wt (ml/g)	Final Volume	Reading	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.306	0.468	03/02/2017	15:27
2	ICB		1	50	50	0.000	0.000	03/02/2017	15:29
3	CCV1	0.50	1	50	50	0.331	0.506	03/02/2017	15:29
4	CCB1		1	50	50	0.000	0.000	03/02/2017	15:29
5	PB97302BL		1	50	50	0.000	0.000	03/02/2017	15:30
6	PB97302BS	0.50	1	50	50	0.307	0.469	03/02/2017	15:30
7	I2058-01		1	50	50	0.078	0.119	03/02/2017	15:30
8	I2058-01DUP		1	50	50	0.078	0.119	03/02/2017	15:31
9	I2058-01MS	0.50	1	50	50	0.370	0.566	03/02/2017	15:32
10	I2058-01MSD	0.50	1	50	50	0.370	0.566	03/02/2017	15:32
11	I2063-01		1	50	50	1.318	2.015	03/02/2017	15:33
12	I2063-01		5	50	50	0.525	0.803	03/02/2017	15:35
13	CCV2	0.50	1	50	50	0.308	0.471	03/02/2017	15:36
14	CCB2		1	50	50	0.000	0.000	03/02/2017	15:37