


7470A

LB86128
INSTRUMENT ID: CVL

Linear

 μ Abs.:

3611E



A= 0.0000e+000

B= 2.7606e-004

C= 1.4494e-002

Rho= 0.9999563

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.016	0.016	6	0.000	6				
0.2	0.200	0.201	0.001	675	0.0 %	675				
2.5	2.500	2.497	-0.003	8994	0.0 %	8994				
5.0	5.000	4.942	-0.058	17851	0.0 %	17851				
7.5	7.500	7.558	0.058	27325	0.0 %	27325				
10.0	10.000	9.985	-0.015	36118	0.0 %	36118				

LB86128 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Mar 2017 14:15:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	6	-	0.0000	7470A	PPB	02 Mar 2017 14:19:31	S	Std
0.2 - 1	50.2	675	-	0.2000	7470A	PPB	02 Mar 2017 14:21:39	S	Std
2.5 - 1	52.5	8994	-	2.5000	7470A	PPB	02 Mar 2017 14:23:48	S	Std
5.0 - 1	55	17851	-	5.0000	7470A	PPB	02 Mar 2017 14:25:56	S	Std
7.5 - 1	57.5	27325	-	7.5000	7470A	PPB	02 Mar 2017 14:28:04	S	Std
10.0 - 1	510	36118	-	10.0000	7470A	PPB	02 Mar 2017 14:32:51	S	Std
ICV56 - 1	ICV56	14218	3.9395	-7470A	PPB	02 Mar 2017 14:35:29	U	SMPL	
ICB56 - 1	ICB56	-74	-0.0059	-7470A	PPB	02 Mar 2017 14:37:35	U	SMPL	
CCV24 - 1	CCV24	18447	5.1070	-7470A	PPB	02 Mar 2017 14:39:41	U	SMPL	
CCB24 - 1	CCB24	-73	-0.0057	-7470A	PPB	02 Mar 2017 14:41:48	U	SMPL	
CRA - 1	CRA	644	0.1923	-7470A	PPB	02 Mar 2017 14:43:54	U	SMPL	
HighStd - 1	HighStd	35858	9.9135	-7470A	PPB	02 Mar 2017 14:46:01	U	SMPL	
ChkStd - 1	ChkStd	25890	7.1617	-7470A	PPB	02 Mar 2017 14:48:08	U	SMPL	
PB97274BL - 1	PBW	-104	-0.0142	-7470A	PPB	02 Mar 2017 14:50:16	U	SMPL	
PB97274BS - 1	LCSW	14453	4.0044	-7470A	PPB	02 Mar 2017 14:52:23	U	SMPL	
PB97274TB - 1	PB97240TB	-57	-0.0012	-7470A	PPB	02 Mar 2017 14:54:31	U	SMPL	
I2006-09 - 1	CPSIDW-1	-13	0.0109	-7470A	PPB	02 Mar 2017 14:56:39	U	SMPL	
I2006-09DUP - 1	CPSIDW-1DUP	-10	0.0117	-7470A	PPB	02 Mar 2017 14:58:47	U	SMPL	
I2006-09MS - 1	CPSIDW-1MS	13734	3.8059	-7470A	PPB	02 Mar 2017 15:00:56	U	SMPL	
I2006-09MSD - 1	CPSIDW-1MSD	13001	3.6036	-7470A	PPB	02 Mar 2017 15:03:02	U	SMPL	
CCV25 - 1	CCV25	18626	5.1564	-7470A	PPB	02 Mar 2017 15:05:09	U	SMPL	
CCB25 - 1	CCB25	-74	-0.0059	-7470A	PPB	02 Mar 2017 15:07:16	U	SMPL	
PB97275BL - 1	PBW	-4	0.0134	-7470A	PPB	02 Mar 2017 15:09:23	U	SMPL	
PB97275BS - 1	LCSW	13850	3.8379	-7470A	PPB	02 Mar 2017 15:11:29	U	SMPL	
PB97275TB - 1	PB97248TB	-60	-0.0021	-7470A	PPB	02 Mar 2017 15:13:36	U	SMPL	
I2015-04 - 1	37A-9-11	200	0.0697	-7470A	PPB	02 Mar 2017 15:20:34	U	SMPL	
I2015-04DUP - 1	37A-9-11DUP	181	0.0645	-7470A	PPB	02 Mar 2017 15:22:41	U	SMPL	
I2015-04MS - 1	37A-9-11MS	14977	4.1491	-7470A	PPB	02 Mar 2017 15:24:48	U	SMPL	
I2015-04MSD - 1	37A-9-11MSD	14218	3.9395	-7470A	PPB	02 Mar 2017 15:26:57	U	SMPL	
PB97277BL - 1	PBW	39	0.0253	-7470A	PPB	02 Mar 2017 15:29:05	U	SMPL	
PB97277BS - 1	LCSW	13861	3.8410	-7470A	PPB	02 Mar 2017 15:31:13	U	SMPL	
I2006-01 - 1	C19013051	183	0.0650	-7470A	PPB	02 Mar 2017 15:33:20	U	SMPL	
CCV26 - 1	CCV26	18370	5.0857	-7470A	PPB	02 Mar 2017 15:35:27	U	SMPL	
CCB26 - 1	CCB26	-60	-0.0021	-7470A	PPB	02 Mar 2017 15:37:34	U	SMPL	
I2006-01DUP - 1	C19013051DUP	200	0.0697	-7470A	PPB	02 Mar 2017 15:39:40	U	SMPL	
I2006-03 - 1	I2006-01MS	14698	4.0720	-7470A	PPB	02 Mar 2017 15:41:47	U	SMPL	
I2006-04 - 1	I2006-01MSD	13901	3.8520	-7470A	PPB	02 Mar 2017 15:43:54	U	SMPL	
I2006-05 - 1	CPS043051	-33	0.0054	-7470A	PPB	02 Mar 2017 15:46:01	U	SMPL	
I2006-06 - 1	CPS043055	-5	0.0131	-7470A	PPB	02 Mar 2017 15:48:09	U	SMPL	
I2006-07 - 1	CPS073051	37	0.0247	-7470A	PPB	02 Mar 2017 15:50:17	U	SMPL	
I2006-08 - 1	CPS073052	45	0.0269	-7470A	PPB	02 Mar 2017 15:52:25	U	SMPL	
I2006-10 - 1	C19013051	29	0.0225	-7470A	PPB	02 Mar 2017 15:54:33	U	SMPL	
I2006-12 - 1	I2006-10MS	14305	3.9636	-7470A	PPB	02 Mar 2017 15:56:42	U	SMPL	
I2006-13 - 1	I2006-10MSD	13675	3.7896	-7470A	PPB	02 Mar 2017 15:58:48	U	SMPL	
CCV27 - 1	CCV27	18599	5.1490	-7470A	PPB	02 Mar 2017 16:00:55	U	SMPL	
CCB27 - 1	CCB27	-55	-0.0007	-7470A	PPB	02 Mar 2017 16:03:02	U	SMPL	
I2006-14 - 1	CPS043051	-15	0.0104	-7470A	PPB	02 Mar 2017 16:05:09	U	SMPL	
I2006-15 - 1	CPS073051	-2	0.0139	-7470A	PPB	02 Mar 2017 16:07:16	U	SMPL	
I2006-16 - 1	CPS073052	-4	0.0134	-7470A	PPB	02 Mar 2017 16:09:23	U	SMPL	
I2052-01 - 1	CPS093051	72	0.0344	-7470A	PPB	02 Mar 2017 16:11:30	U	SMPL	
I2052-02 - 1	CPS123051	61	0.0313	-7470A	PPB	02 Mar 2017 16:13:37	U	SMPL	
I2052-03 - 1	CPS133051	77	0.0358	-7470A	PPB	02 Mar 2017 16:15:45	U	SMPL	
I2052-04 - 1	CPS103051	30	0.0228	-7470A	PPB	02 Mar 2017 16:17:52	U	SMPL	
I2052-05 - 1	CPS113051	28	0.0222	-7470A	PPB	02 Mar 2017 16:20:01	U	SMPL	
I2052-06 - 1	CPS093051	51	0.0286	-7470A	PPB	02 Mar 2017 16:22:09	U	SMPL	
I2052-07 - 1	CPS123051	19	0.0197	-7470A	PPB	02 Mar 2017 16:24:16	U	SMPL	
CCV28 - 1	CCV28	17633	4.8823	-7470A	PPB	02 Mar 2017 16:26:23	U	SMPL	
CCB28 - 1	CCB28	-47	0.0015	-7470A	PPB	02 Mar 2017 16:28:30	U	SMPL	
I2052-08 - 1	CPS133051	116	0.0465	-7470A	PPB	02 Mar 2017 16:30:37	U	SMPL	
I2052-09 - 1	CPS103051	-12	0.0112	-7470A	PPB	02 Mar 2017 16:32:44	U	SMPL	
I2052-10 - 1	CPS113051	-6	0.0128	-7470A	PPB	02 Mar 2017 16:34:51	U	SMPL	
I2006-09LX5 - 1		-4	0.0134	-7470A	PPB	02 Mar 2017 16:36:58	U	SMPL	
I2006-09A - 1		14253	3.9492	-7470A	PPB	02 Mar 2017 16:39:06	U	SMPL	
I2015-04LX5 - 1		-14	0.0106	-7470A	PPB	02 Mar 2017 16:41:14	U	SMPL	
I2015-04A - 1		14311	3.9652	-7470A	PPB	02 Mar 2017 16:43:21	U	SMPL	
I2006-01LX5 - 1		-44	0.0023	-7470A	PPB	02 Mar 2017 16:45:29	U	SMPL	
I2006-01A - 1		14589	4.0420	-7470A	PPB	02 Mar 2017 16:47:38	U	SMPL	
CCV29 - 1	CCV29	17807	4.9303	-7470A	PPB	02 Mar 2017 16:49:45	U	SMPL	
CCB29 - 1	CCB29	-69	-0.0046	-7470A	PPB	02 Mar 2017 16:51:52	U	SMPL	