



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86132

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/02/2017	17:25
2	CAL2	1	Water			7.02	03/02/2017	17:26
3	CAL3	1	Water			10.04	03/02/2017	17:27
4	ICV	1	Water			7.01	03/02/2017	17:28
5	CCV1	1	Water			2.01	03/02/2017	17:30
6	I2089-03	1	Solid	20.02	20	8.24	03/02/2017	17:31
7	I2089-03DUP	1	Solid	20.04	20	8.25	03/02/2017	17:32
8	I2089-04	1	Solid	20.05	20	7.61	03/02/2017	17:33
9	CCV2	1	Water			12.02	03/02/2017	17:34

1-1386132-

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I2089

WorkList ID : 96117

Date : 3/2/2017 5:20:59 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2089-03	Corrosivity	Cool 4 deg C	ENTA04	J23	WC-B62-B64-001	03/02/2017	9045C
	Solid	I2089-04	Corrosivity	Cool 4 deg C	ENTA04	J23	WC-B65-B66-001	03/02/2017	9045C

Date/Time 03/02/17 5:20PM
 Received by: 31
 Relinquished by: 68

Date/Time 03/02/17 6:30PM
 Received by: EP
 Relinquished by: SD