



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86143

Temp.: 18.6 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/03/2017	13:00
2	CAL2	1	Water			7.02	03/03/2017	13:01
3	CAL3	1	Water			10.03	03/03/2017	13:02
4	ICV	1	Water			7.01	03/03/2017	13:03
5	CCV1	1	Water			2.02	03/03/2017	13:05
6	I2100-01	1	Water			7.34	03/03/2017	13:06
7	I2100-01DUP	1	Water			7.35	03/03/2017	13:07
8	CCV2	1	Water			12.01	03/03/2017	13:10

WORKLIST(Hardcopy Internal Chain)

1386143

WorkList Name: PH I2100

WorkList ID: 96151

Date: 3/3/2017 12:44:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/03/2017	Water	I2100-01	pH	Cool 4 deg C	PSEG01	G31	FRAC-TANK-252657	03/03/2017	9045C

Date/Time 03/03/17 12:44 PM
 Received by: JP
 Relinquished by: JP

Date/Time 03/03/17 2:00 PM
 Received by: JP
 Relinquished by: JP