


245.1

LB86148
INSTRUMENT ID: CV 4

Linear

 μ Abs.:

34111



A= 0.0000e+000

B= 2.9364e-004

C= 1.4883e-002

Rho= 0.9999697

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.019	0.019	14	0.000	14				
0.2	0.200	0.190	-0.010	598	0.0 %	598				
2.5	2.500	2.527	0.027	8555	0.0 %	8555				
5.0	5.000	4.950	-0.050	16807	0.0 %	16807				
7.5	7.500	7.482	-0.018	25431	0.0 %	25431				
10.0	10.000	10.031	0.031	34111	0.0 %	34111				

LB86148 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 03 Mar 2017 14:42:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	14	-	0.0000	245.1	PPB	03 Mar 2017 14:48:14	S	Std
0.2 - 1	50.2	598	-	0.2000	245.1	PPB	03 Mar 2017 14:50:22	S	Std
2.5 - 1	52.5	8555	-	2.5000	245.1	PPB	03 Mar 2017 14:52:31	S	Std
5.0 - 1	55	16807	-	5.0000	245.1	PPB	03 Mar 2017 14:54:39	S	Std
7.5 - 1	57.5	25431	-	7.5000	245.1	PPB	03 Mar 2017 14:56:47	S	Std
10.0 - 1	510	34111	-	10.0000	245.1	PPB	03 Mar 2017 14:58:54	S	Std
ICV57 - 1	ICV57	13480	3.9731	-245.1	PPB	PPB	03 Mar 2017 15:01:33	U	SMPL
ICB57 - 1	ICB57	-49	0.0005	-245.1	PPB	PPB	03 Mar 2017 15:03:39	U	SMPL
CCV30 - 1	CCV30	16878	4.9709	-245.1	PPB	PPB	03 Mar 2017 15:05:45	U	SMPL
CCB30 - 1	CCB30	-46	0.0014	-245.1	PPB	PPB	03 Mar 2017 15:07:51	U	SMPL
CRA - 1	CRA	584	0.1864	-245.1	PPB	PPB	03 Mar 2017 15:09:58	U	SMPL
HighStd - 1	HighStd	34545	10.1586	-245.1	PPB	PPB	03 Mar 2017 15:12:05	U	SMPL
ChkStd - 1	ChkStd	24057	7.0789	-245.1	PPB	PPB	03 Mar 2017 15:14:12	U	SMPL
PB97320BL - 1	PBW	-97	-0.0136	-245.1	PPB	PPB	03 Mar 2017 15:16:19	U	SMPL
PB97320BS - 1	LCSW	13419	3.9552	-245.1	PPB	PPB	03 Mar 2017 15:18:27	U	SMPL
I2081-01 - 1	022717-PRRT-F-D-M	32	0.0243	-245.1	PPB	PPB	03 Mar 2017 15:20:34	U	SMPL
I2081-02 - 1	022717-PRRT-F-U-M	5	0.0164	-245.1	PPB	PPB	03 Mar 2017 15:22:42	U	SMPL
I2082-01 - 1	030117-PRRT-F-D-M	68	0.0349	-245.1	PPB	PPB	03 Mar 2017 15:24:50	U	SMPL
I2082-02 - 1	030117-PRRT-F-U-M	-10	0.0119	-245.1	PPB	PPB	03 Mar 2017 15:26:59	U	SMPL
I2083-01 - 1	022817-PRRT-F-D-M	2	0.0155	-245.1	PPB	PPB	03 Mar 2017 15:29:06	U	SMPL
CCV31 - 1	CCV31	16714	4.9227	-245.1	PPB	PPB	03 Mar 2017 15:31:13	U	SMPL
CCB31 - 1	CCB31	-44	0.0020	-245.1	PPB	PPB	03 Mar 2017 15:33:20	U	SMPL
I2083-02 - 1	022817-PRRT-F-U-M	6	0.0166	-245.1	PPB	PPB	03 Mar 2017 15:35:26	U	SMPL
I2083-02DUP - 1	022817-PRRT-F-U-MDUP	5	0.0164	-245.1	PPB	PPB	03 Mar 2017 15:37:33	U	SMPL
I2083-02MS - 1	022817-PRRT-F-U-MMS	15912	4.6872	-245.1	PPB	PPB	03 Mar 2017 15:39:40	U	SMPL
I2083-02MSD - 1	022817-PRRT-F-U-MMSD	15766	4.6444	-245.1	PPB	PPB	03 Mar 2017 15:41:47	U	SMPL
I2083-02LX5 - 1		-49	0.0005	-245.1	PPB	PPB	03 Mar 2017 15:43:55	U	SMPL
I2083-02A - 1		13655	4.0245	-245.1	PPB	PPB	03 Mar 2017 15:46:02	U	SMPL
CCV32 - 1	CCV32	16798	4.9474	-245.1	PPB	PPB	03 Mar 2017 15:48:11	U	SMPL
CCB32 - 1	CCB32	-43	0.0023	-245.1	PPB	PPB	03 Mar 2017 15:50:19	U	SMPL