

7470A


LB86164

INSTRUMENT ID: CV 1

Linear

 μ Abs.:

35773



A= 0.0000e+000

B= 2.7833e-004

C= 7.6129e-003

Rho= 0.9999275

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	3	0.000	3				
0.2	0.200	0.225	0.025	780	0.0 %	780				
2.5	2.500	2.420	-0.080	8669	0.0 %	8669				
5.0	5.000	5.031	0.031	18047	0.0 %	18047				
7.5	7.500	7.551	0.051	27103	0.0 %	27103				
10.0	10.000	9.964	-0.036	35773	0.0 %	35773				

LB86164 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Mar 2017 10:47:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	3	-	0.00007470A	PPB		06 Mar 2017 11:00:26	S	Std
0.2 - 1	SO-2	780	-	0.20007470A	PPB		06 Mar 2017 11:02:34	S	Std
2.5 - 1	SO-5	8669	-	2.50007470A	PPB		06 Mar 2017 11:04:42	S	Std
5.0 - 1	SO	18047	-	5.00007470A	PPB		06 Mar 2017 11:06:51	S	Std
7.5 - 1	SO-5	27103	-	7.50007470A	PPB		06 Mar 2017 11:08:58	S	Std
10.0 - 1	SO	35773	-	10.00007470A	PPB		06 Mar 2017 11:11:05	S	Std
ICV58 - 1	ICV58	13994	3.9026	-7470A	PPB		06 Mar 2017 11:13:45	U	SMPL
ICB58 - 1	ICB58	-92	-0.0180	-7470A	PPB		06 Mar 2017 11:15:51	U	SMPL
CCV35 - 1	CCV35	17976	5.0109	-7470A	PPB		06 Mar 2017 11:17:57	U	SMPL
CCB35 - 1	CCB35	-96	-0.0191	-7470A	PPB		06 Mar 2017 11:20:04	U	SMPL
CRA - 1	CRA	774	0.2230	-7470A	PPB		06 Mar 2017 11:22:11	U	SMPL
HighStd - 1	HighStd	35227	9.8124	-7470A	PPB		06 Mar 2017 11:24:17	U	SMPL
ChkStd - 1	ChkStd	25309	7.0519	-7470A	PPB		06 Mar 2017 11:26:24	U	SMPL
PB97335BL - 1	PBW	-125	-0.0272	-7470A	PPB		06 Mar 2017 11:28:32	U	SMPL
PB97335BS - 1	LCSW	13604	3.7941	-7470A	PPB		06 Mar 2017 11:30:39	U	SMPL
PB97335TB - 1	PB97295TB	-82	-0.0152	-7470A	PPB		06 Mar 2017 11:32:47	U	SMPL
I2089-03 - 1	WC-B62-B64-001	19	0.0129	-7470A	PPB		06 Mar 2017 11:34:55	U	SMPL
I2089-04 - 1	WC-B65-B66-001	-15	0.0034	-7470A	PPB		06 Mar 2017 11:37:03	U	SMPL
I2089-04DUP - 1	WC-B65-B66-001DUP	-3	0.0068	-7470A	PPB		06 Mar 2017 11:39:12	U	SMPL
I2089-04MS - 1	WC-B65-B66-001MS	13452	3.7517	-7470A	PPB		06 Mar 2017 11:41:18	U	SMPL
CCV36 - 1	CCV36	17972	5.0098	-7470A	PPB		06 Mar 2017 11:43:25	U	SMPL
CCB36 - 1	CCB36	-97	-0.0194	-7470A	PPB		06 Mar 2017 11:45:32	U	SMPL
I2089-04MSD - 1	WC-B65-B66-001MSD	14462	4.0329	-7470A	PPB		06 Mar 2017 11:47:39	U	SMPL
PB97336BL - 1	PBW	-102	-0.0208	-7470A	PPB		06 Mar 2017 11:49:46	U	SMPL
PB97336BS - 1	LCSW	13710	3.8236	-7470A	PPB		06 Mar 2017 11:51:53	U	SMPL
I2100-01 - 1	FRAC-TANK-252657	161	0.0524	-7470A	PPB		06 Mar 2017 11:54:00	U	SMPL
I2100-01DUP - 1	FRAC-TANK-252657DUP	258	0.0794	-7470A	PPB		06 Mar 2017 11:56:07	U	SMPL
I2100-01MS - 1	FRAC-TANK-252657MS	15976	4.4543	-7470A	PPB		06 Mar 2017 11:58:44	U	SMPL
I2100-01MSD - 1	FRAC-TANK-252657MSD	16001	4.4612	-7470A	PPB		06 Mar 2017 12:00:52	U	SMPL
CCV37 - 1	CCV37	17998	5.0170	-7470A	PPB		06 Mar 2017 12:03:00	U	SMPL
CCB37 - 1	CCB37	-82	-0.0152	-7470A	PPB		06 Mar 2017 12:05:09	U	SMPL
I2089-04LX5 - 1		-25	0.0007	-7470A	PPB		06 Mar 2017 12:09:22	U	SMPL
I2089-04A - 1		13936	3.8865	-7470A	PPB		06 Mar 2017 12:11:29	U	SMPL
I2100-01LX5 - 1		-22	0.0015	-7470A	PPB		06 Mar 2017 12:13:36	U	SMPL
I2100-01A - 1		13668	3.8119	-7470A	PPB		06 Mar 2017 12:18:13	U	SMPL
CCV38 - 1	CCV38	17867	4.9806	-7470A	PPB		06 Mar 2017 12:20:19	U	SMPL
CCB38 - 1	CCB38	-122	-0.0263	-7470A	PPB		06 Mar 2017 12:22:26	U	SMPL