



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB86165

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	03/06/2017	11:15
2	I2100-02		1	>100.0	>212.0	03/06/2017	11:45
3	I2100-02DUP		1	>100.0	>212.0	03/06/2017	12:16

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 86165

WorkList Name : FP3617

WorkList ID : 96190

Date : 3/6/2017 10:49:32 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/08/2017	Water	12100-02 K	Flash Point	Cool 4 deg C	PSEG01	G31	FRAC-TANK-252657	03/03/2017	1010A

Date/Time 03/06/17 10:50 AM
 Received by: KS
 Relinquished by: GP

Date/Time 03/06/17 12:25 PM
 Received by: GP
 Relinquished by: KS