

7471A


LB 86182

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4564E



A= 0.0000e+000

B= 2.1927e-004

C= -1.7216e-002

Rho= 0.9999705

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.012	-0.012	24	0.000	24				
0.2	0.200	0.172	-0.028	862	0.0 %	862				
2.5	2.500	2.535	0.035	11638	0.0 %	11638				
5.0	5.000	5.042	0.042	23075	0.0 %	23075				
7.5	7.500	7.472	-0.028	34153	0.0 %	34153				
10.0	10.000	9.992	-0.008	45646	0.0 %	45646				

LB86182 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 06 Mar 2017 14:49:14

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SC	24	-	0.0000	7471A	PPB	06 Mar 2017 14:50:52	S	Std
0.2 - 1	SC-2	862	-	0.2000	7471A	PPB	06 Mar 2017 14:53:09	S	Std
2.5 - 1	SC-5	11638	-	2.5000	7471A	PPB	06 Mar 2017 14:55:27	S	Std
5.0 - 1	SC-5	23075	-	5.0000	7471A	PPB	06 Mar 2017 14:57:44	S	Std
7.5 - 1	SC-5	34153	-	7.5000	7471A	PPB	06 Mar 2017 15:00:02	S	Std
10.0 - 1	SC-5	45646	-	10.0000	7471A	PPB	06 Mar 2017 15:02:18	S	Std
ICV59 - 1	ICV59	17192	3.7525	-7471A	PPB	06 Mar 2017 15:05:01	U	SMPL	
ICB59 - 1	ICB59	-70	-0.0326	-7471A	PPB	06 Mar 2017 15:07:16	U	SMPL	
CCV39 - 1	CCV39	22976	5.0207	-7471A	PPB	06 Mar 2017 15:09:32	U	SMPL	
CCB39 - 1	CCB39	-72	-0.0330	-7471A	PPB	06 Mar 2017 15:11:48	U	SMPL	
CRA - 1	CRA	821	0.1628	-7471A	PPB	06 Mar 2017 15:14:03	U	SMPL	
HighStd - 1	HighStd	44682	9.7802	-7471A	PPB	06 Mar 2017 15:16:19	U	SMPL	
ChkStd - 1	ChkStd	32939	7.2053	-7471A	PPB	06 Mar 2017 15:18:35	U	SMPL	
PB97368BL - 1	PBS	-96	-0.0383	-7471A	PPB	06 Mar 2017 15:20:51	U	SMPL	
PB97368BS - 1	LCSS	15045	3.2817	-7471A	PPB	06 Mar 2017 15:23:08	U	SMPL	
I2061-01 - 1	IMPACT	1394	0.2884	-7471A	PPB	06 Mar 2017 15:25:25	U	SMPL	
I2061-02 - 1	IMPACT-2-FROMDUMPSTER	1024	0.2073	-7471A	PPB	06 Mar 2017 15:27:42	U	SMPL	
I2061-03 - 1	IMPACT-3	3532	0.7572	-7471A	PPB	06 Mar 2017 15:30:00	U	SMPL	
I2076-02 - 1	39A-7-9	191	0.0247	-7471A	PPB	06 Mar 2017 15:32:17	U	SMPL	
I2078-01 - 1	C-1	498	0.0920	-7471A	PPB	06 Mar 2017 15:34:33	U	SMPL	
CCV40 - 1	CCV40	22862	4.9957	-7471A	PPB	06 Mar 2017 15:36:49	U	SMPL	
CCB40 - 1	CCB40	-36	-0.0251	-7471A	PPB	06 Mar 2017 15:39:06	U	SMPL	
I2078-03 - 1	C-2	867	0.1729	-7471A	PPB	06 Mar 2017 15:41:22	U	SMPL	
I2078-05 - 1	C-3	600	0.1143	-7471A	PPB	06 Mar 2017 15:43:38	U	SMPL	
I2078-07 - 1	C-4	445	0.0804	-7471A	PPB	06 Mar 2017 15:45:54	U	SMPL	
I2078-09 - 1	C-5	277	0.0435	-7471A	PPB	06 Mar 2017 15:48:10	U	SMPL	
I2078-11 - 1	C-6	698	0.1358	-7471A	PPB	06 Mar 2017 15:50:27	U	SMPL	
I2078-13 - 1	C-7	590	0.1122	-7471A	PPB	06 Mar 2017 15:52:44	U	SMPL	
I2078-15 - 1	C-8	623	0.1194	-7471A	PPB	06 Mar 2017 15:55:01	U	SMPL	
I2087-01 - 1	FK-BPM-02-002-ABCD	333	0.0558	-7471A	PPB	06 Mar 2017 15:57:19	U	SMPL	
I2087-01DUP - 1	FK-BPM-02-002-ABCD DUP	354	0.0604	-7471A	PPB	06 Mar 2017 15:59:37	U	SMPL	
I2087-01MS - 1	FK-BPM-02-002-ABCDMS	16638	3.6310	-7471A	PPB	06 Mar 2017 16:01:53	U	SMPL	
CCV41 - 1	CCV41	23116	5.0514	-7471A	PPB	06 Mar 2017 16:04:09	U	SMPL	
CCB41 - 1	CCB41	-24	-0.0225	-7471A	PPB	06 Mar 2017 16:06:26	U	SMPL	
I2087-01MSD - 1	FK-BPM-02-002-ABCDMSD	16399	3.5786	-7471A	PPB	06 Mar 2017 16:08:42	U	SMPL	
I2088-01 - 1	FK-GF-02-004-A	1913	0.4022	-7471A	PPB	06 Mar 2017 16:10:59	U	SMPL	
I2088-03 - 1	FK-GF-02-004-B	1406	0.2911	-7471A	PPB	06 Mar 2017 16:13:15	U	SMPL	
2091-01 - 1	RF-1	630	0.1209	-7471A	PPB	06 Mar 2017 16:15:31	U	SMPL	
2099-02 - 1	39B-9-11	503	0.0931	-7471A	PPB	06 Mar 2017 16:17:47	U	SMPL	
2099-04 - 1	39C-9-11	138	0.0130	-7471A	PPB	06 Mar 2017 16:20:04	U	SMPL	
2087-01LX5 - 1		67	-0.0025	-7471A	PPB	06 Mar 2017 16:22:21	U	SMPL	
2087-01A - 1		16426	3.5845	-7471A	PPB	06 Mar 2017 16:24:39	U	SMPL	
CCV42 - 1	CCV42	23150	5.0589	-7471A	PPB	06 Mar 2017 16:26:56	U	SMPL	
CCB42 - 1	CCB42	-23	-0.0223	-7471A	PPB	06 Mar 2017 16:29:13	U	SMPL	