



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86195

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/03/2017	16:30
2	CAL2	1	Water			7.02	03/03/2017	16:31
3	CAL3	1	Water			10.03	03/03/2017	16:32
4	ICV	1	Water			6.99	03/03/2017	16:33
5	CCV1	1	Water			2.01	03/03/2017	16:35
6	I2103-01	1	Solid	20.03	20	8.57	03/03/2017	16:36
7	I2103-01DUP	1	Solid	20.05	20	8.58	03/03/2017	16:37
8	CCV2	1	Water			12.01	03/03/2017	16:38

WORKLIST(Hardcopy Internal Chain)

1886195
03/03/17

WorkList Name : CORROSIONITY I2103

WorkList ID : 96252

Date : 3/2/2017 4:10:27 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2103-01	Corrosivity	Cool 4 deg C	ENTA04	K21	WC-B46-B47-001	03/02/2017	9045C

03/03/2017
Date/Time 03/03/17 4:10 PM
Received by: TH
Relinquished by: CR

01/03/17
Date/Time 03/03/17 8:30 PM
Received by: CR
Relinquished by: JP