


7470A

LB 86196
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35217



A= 0.0000e+000

B= 2.8530e-004

C= 2.9828e-002

Rho= 0.9998049

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.029	0.029	-2	0.000	-2				
0.2	0.200	0.227	0.027	692	0.0 %	692				
2.5	2.500	2.536	0.036	8786	0.0 %	8786				
5.0	5.000	4.850	-0.150	16894	0.0 %	16894				
7.5	7.500	7.480	-0.020	26114	0.0 %	26114				
10.0	10.000	10.077	0.077	35217	0.0 %	35217				

LB86196 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Mar 2017 10:02:30

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	-2	-	0.0000	7470A	PPB	07 Mar 2017 10:06:49	S	Std
0.2 - 1	50.2	692	-	0.2000	7470A	PPB	07 Mar 2017 10:08:58	S	Std
2.5 - 1	52.5	8786	-	2.5000	7470A	PPB	07 Mar 2017 10:11:06	S	Std
5.0 - 1	55	16894	-	5.0000	7470A	PPB	07 Mar 2017 10:13:14	S	Std
7.5 - 1	57.5	26114	-	7.5000	7470A	PPB	07 Mar 2017 10:15:22	S	Std
10.0 - 1	510	35217	-	10.0000	7470A	PPB	07 Mar 2017 10:19:59	S	Std
ICV60 - 1	ICV60	13634	3.9196	-	7470A	PPB	07 Mar 2017 10:22:39	U	SMPL
ICB60 - 1	ICB60	-70	0.0099	-	7470A	PPB	07 Mar 2017 10:24:45	U	SMPL
CCV43 - 1	CCV43	17078	4.9022	-	7470A	PPB	07 Mar 2017 10:26:51	U	SMPL
CCB43 - 1	CCB43	-77	0.0079	-	7470A	PPB	07 Mar 2017 10:28:58	U	SMPL
CRA - 1	CRA	680	0.2238	-	7470A	PPB	07 Mar 2017 10:31:04	U	SMPL
HighStd - 1	HighStd	34954	10.0022	-	7470A	PPB	07 Mar 2017 10:33:11	U	SMPL
ChkStd - 1	ChkStd	24047	6.8904	-	7470A	PPB	07 Mar 2017 10:35:18	U	SMPL
PB97382BL - 1	PBW	-90	0.0042	-	7470A	PPB	07 Mar 2017 10:37:26	U	SMPL
PB97382BS - 1	LCSW	12839	3.6928	-	7470A	PPB	07 Mar 2017 10:39:34	U	SMPL
PB97382TB - 1	PB97343TB	103	0.0592	-	7470A	PPB	07 Mar 2017 10:41:41	U	SMPL
I2100-02 - 1	FRAC-TANK-252657	-37	0.0193	-	7470A	PPB	07 Mar 2017 10:43:50	U	SMPL
I2103-01 - 1	WC-B46-B47-001	8	0.0321	-	7470A	PPB	07 Mar 2017 10:45:58	U	SMPL
I2103-01DUP - 1	WC-B46-B47-001DUP	-2	0.0293	-	7470A	PPB	07 Mar 2017 10:48:07	U	SMPL
I2103-01MS - 1	WC-B46-B47-001MS	12258	3.5270	-	7470A	PPB	07 Mar 2017 10:50:13	U	SMPL
CCV44 - 1	CCV44	17093	4.9064	-	7470A	PPB	07 Mar 2017 10:52:19	U	SMPL
CCB44 - 1	CCB44	-82	0.0064	-	7470A	PPB	07 Mar 2017 10:54:26	U	SMPL
I2103-01MSD - 1	WC-B46-B47-001MSD	12531	3.6049	-	7470A	PPB	07 Mar 2017 10:56:33	U	SMPL
I2103-01LX5 - 1		-72	0.0093	-	7470A	PPB	07 Mar 2017 10:58:40	U	SMPL
I2103-01A - 1		13628	3.9179	-	7470A	PPB	07 Mar 2017 11:00:47	U	SMPL
CCV45 - 1	CCV45	16830	4.8314	-	7470A	PPB	07 Mar 2017 11:02:54	U	SMPL
CCB45 - 1	CCB45	-77	0.0079	-	7470A	PPB	07 Mar 2017 11:05:01	U	SMPL