



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86209

Temp.: 23.6 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/07/2017	14:10
2	CAL2	1	Water			7.00	03/07/2017	14:11
3	CAL3	1	Water			10.02	03/07/2017	14:12
4	ICV	1	Water			7.01	03/07/2017	14:15
5	CCV1	1	Water			2.01	03/07/2017	14:16
6	I2127-01	1	Solid	20.02	20	5.71	03/07/2017	14:17
7	I2127-01DUP	1	Solid	20.04	20	5.72	03/07/2017	14:18
8	I2127-03	1	Solid	20.02	20	5.86	03/07/2017	14:19
9	I2127-05	1	Solid	20.01	20	6.24	03/07/2017	14:20
10	I2127-07	1	Solid	20.03	20	6.74	03/07/2017	14:21
11	I2127-09	1	Solid	20.05	20	6.82	03/07/2017	14:22
12	I2127-11	1	Solid	20.04	20	7.02	03/07/2017	14:23
13	I2127-13	1	Solid	20.01	20	6.90	03/07/2017	14:25
14	I2127-15	1	Solid	20.04	20	6.85	03/07/2017	14:26
15	I2127-17	1	Solid	20.02	20	6.18	03/07/2017	14:30
16	CCV2	1	Water			12.02	03/07/2017	14:31
17	I2129-01	1	Solid	20.00	20	6.44	03/07/2017	14:32
18	I2129-03	1	Solid	20.03	20	6.36	03/07/2017	14:33
19	I2129-05	1	Solid	20.04	20	6.24	03/07/2017	14:35
20	I2129-07	1	Solid	20.02	20	6.43	03/07/2017	14:36

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water			12.02	03/07/2017	14:37

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I2127

WorkList ID : 96261

Date : 3/7/2017 1:57:22 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2127-01	pH	Cool 4 deg C	SEVE02	K43	D-1	03/06/2017	9045C
	Solid	I2127-03	pH	Cool 4 deg C	SEVE02	K43	D-8	03/06/2017	9045C
	Solid	I2127-05	pH	Cool 4 deg C	SEVE02	K43	D-12	03/06/2017	9045C
	Solid	I2127-07	pH	Cool 4 deg C	SEVE02	K43	D-14	03/06/2017	9045C
	Solid	I2127-09	pH	Cool 4 deg C	SEVE02	K43	D-18	03/06/2017	9045C
	Solid	I2127-11	pH	Cool 4 deg C	SEVE02	K43	D-21	03/06/2017	9045C
	Solid	I2127-13	pH	Cool 4 deg C	SEVE02	K43	D-23	03/06/2017	9045C
	Solid	I2127-15	pH	Cool 4 deg C	SEVE02	K43	D-25	03/06/2017	9045C
	Solid	I2127-17	pH	Cool 4 deg C	SEVE02	K43	D-26	03/06/2017	9045C
	Solid	I2129-01	pH	Cool 4 deg C	SEVE02	K43	D-27	03/06/2017	9045C
	Solid	I2129-03	pH	Cool 4 deg C	SEVE02	K43	D-30	03/06/2017	9045C
	Solid	I2129-05	pH	Cool 4 deg C	SEVE02	K43	D-33	03/06/2017	9045C
	Solid	I2129-07	pH	Cool 4 deg C	SEVE02	K43	D-35	03/06/2017	9045C

Date/Time 03/07/17 1:57 PM
 Received by: JJ
 Relinquished by: SR

Date/Time 03/07/17 6:00 PM
 Received by: SR
 Relinquished by: JJ