


245.1

L386218
INSTRUMENT ID: CV1

Linear

 μ Abs.:

33255



A= 0.0000e+000

B= 3.0060e-004

C= 5.6747e-003

Rho= 0.9999985

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.007	0.007	3	0.000	3				
0.2	0.200	0.191	-0.009	618	0.0 %	618				
2.5	2.500	2.508	0.008	8323	0.0 %	8323				
5.0	5.000	4.992	-0.008	16589	0.0 %	16589				
7.5	7.500	7.500	-0.000	24930	0.0 %	24930				
10.0	10.000	10.002	0.002	33255	0.0 %	33255				

LB86218 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 08 Mar 2017 10:00:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SC	3	-	0.0000	245.1	PPB	08 Mar 2017 10:07:20	S	Std
0.2 - 1	50.2	618	-	0.2000	245.1	PPB	08 Mar 2017 10:09:28	S	Std
2.5 - 1	52.5	8323	-	2.5000	245.1	PPB	08 Mar 2017 10:11:37	S	Std
5.0 - 1	55	16589	-	5.0000	245.1	PPB	08 Mar 2017 10:13:45	S	Std
7.5 - 1	57.5	24930	-	7.5000	245.1	PPB	08 Mar 2017 10:18:24	S	Std
10.0 - 1	510	33255	-	10.0000	245.1	PPB	08 Mar 2017 10:20:31	S	Std
ICV61 - 1	ICV61	13126	3.9514	-245.1	PPB	PPB	08 Mar 2017 10:23:05	U	SMPL
ICB61 - 1	ICB61	-53	-0.0103	-245.1	PPB	PPB	08 Mar 2017 10:25:11	U	SMPL
CCV50 - 1	CCV50	16469	4.9563	-245.1	PPB	PPB	08 Mar 2017 10:27:18	U	SMPL
CCB50 - 1	CCB50	-48	-0.0088	-245.1	PPB	PPB	08 Mar 2017 10:29:24	U	SMPL
CRA - 1	CRA	610	0.1890	-245.1	PPB	PPB	08 Mar 2017 10:31:31	U	SMPL
HighStd - 1	HighStd	33788	10.1625	-245.1	PPB	PPB	08 Mar 2017 10:33:37	U	SMPL
ChkStd - 1	ChkStd	23042	6.9322	-245.1	PPB	PPB	08 Mar 2017 10:35:44	U	SMPL
PB97400BL - 1	PBW	-85	-0.0199	-245.1	PPB	PPB	08 Mar 2017 10:37:52	U	SMPL
PB97400BS - 1	LCSW	12641	3.8056	-245.1	PPB	PPB	08 Mar 2017 10:40:00	U	SMPL
I2123-01 - 1	030217-PRRT-E-D-M	-21	-0.0006	-245.1	PPB	PPB	08 Mar 2017 10:42:08	U	SMPL
I2123-02 - 1	030217-PRRT-E-U-M	7	0.0078	-245.1	PPB	PPB	08 Mar 2017 10:44:16	U	SMPL
I2124-01 - 1	030317-PRRT-E-D-M	1	0.0060	-245.1	PPB	PPB	08 Mar 2017 10:46:24	U	SMPL
I2124-02 - 1	030317-PRRT-E-U-M	-1	0.0054	-245.1	PPB	PPB	08 Mar 2017 10:48:33	U	SMPL
I2124-02DUP - 1	030317-PRRT-E-U-MDUP	1	0.0060	-245.1	PPB	PPB	08 Mar 2017 10:50:40	U	SMPL
CCV51 - 1	CCV51	16276	4.8983	-245.1	PPB	PPB	08 Mar 2017 10:52:46	U	SMPL
CCB51 - 1	CCB51	-57	-0.0115	-245.1	PPB	PPB	08 Mar 2017 10:54:53	U	SMPL
I2124-02MS - 1	030317-PRRT-E-U-MMS	13774	4.1462	-245.1	PPB	PPB	08 Mar 2017 10:57:00	U	SMPL
I2124-02MSD - 1	030317-PRRT-E-U-MMSD	12783	3.8483	-245.1	PPB	PPB	08 Mar 2017 10:59:07	U	SMPL
I2124-02LX5 - 1		-42	-0.0070	-245.1	PPB	PPB	08 Mar 2017 11:01:14	U	SMPL
I2124-02A - 1		13050	3.9286	-245.1	PPB	PPB	08 Mar 2017 11:03:20	U	SMPL
CCV52 - 1	CCV52	16306	4.9073	-245.1	PPB	PPB	08 Mar 2017 11:05:28	U	SMPL
CCB52 - 1	CCB52	-54	-0.0106	-245.1	PPB	PPB	08 Mar 2017 11:07:35	U	SMPL