


7471A

LB 86220
INSTRUMENT ID: CV2

Linear

 μ Abs.:

44322



A= 0.0000e+000

B= 2.2673e-004

C= 8.3217e-003

Rho= 0.9998047

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.041	0.041	145	0.000	145				
0.2	0.200	0.206	0.006	870	0.0 %	870				
2.5	2.500	2.421	-0.079	10641	0.0 %	10641				
5.0	5.000	5.088	0.088	22406	0.0 %	22406				
7.5	7.500	7.387	-0.113	32543	0.0 %	32543				
10.0	10.000	10.057	0.057	44322	0.0 %	44322				

LB86220 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 08 Mar 2017 11:16:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	145	-	0.00007471A	PPB		08 Mar 2017 11:19:49	S	Std
0.2 - 1	50.2	870	-	0.20007471A	PPB		08 Mar 2017 11:22:06	S	Std
2.5 - 1	52.5	10641	-	2.50007471A	PPB		08 Mar 2017 11:24:24	S	Std
5.0 - 1	55	22406	-	5.00007471A	PPB		08 Mar 2017 11:26:41	S	Std
7.5 - 1	57.5	32543	-	7.50007471A	PPB		08 Mar 2017 11:28:58	S	Std
10.0 - 1	59	44322	-	10.00007471A	PPB		08 Mar 2017 11:33:51	S	Std
ICV62 - 1	ICV62	16476	3.7439	-7471A	PPB		08 Mar 2017 11:37:14	U	SMPL
ICB62 - 1	ICB62	87	0.0280	-7471A	PPB		08 Mar 2017 11:39:30	U	SMPL
CCV53 - 1	CCV53	22411	5.0895	-7471A	PPB		08 Mar 2017 11:41:46	U	SMPL
CCB53 - 1	CCB53	87	0.0280	-7471A	PPB		08 Mar 2017 11:44:01	U	SMPL
CRA - 1	CRA	870	0.2056	-7471A	PPB		08 Mar 2017 11:46:17	U	SMPL
HighStd - 1	HighStd	44994	10.2096	-7471A	PPB		08 Mar 2017 11:48:33	U	SMPL
ChkStd - 1	ChkStd	30223	6.8607	-7471A	PPB		08 Mar 2017 11:50:49	U	SMPL
PB97415BL - 1	PBS	205	0.0548	-7471A	PPB		08 Mar 2017 11:56:38	U	SMPL
PB97415BS - 1	LCSS	15542	3.5321	-7471A	PPB		08 Mar 2017 11:58:54	U	SMPL
I2106-03 - 1	HY-SB414	22900	5.2004	-7471A	PPB		08 Mar 2017 12:01:10	U	SMPL
I2106-04 - 1	HY-SB413A	456	0.1117	-7471A	PPB		08 Mar 2017 12:03:28	U	SMPL
I2106-05 - 1	HY-SB413B	4610	1.0535	-7471A	PPB		08 Mar 2017 12:05:45	U	SMPL
I2126-01 - 1	E2-Z7-NSW	5271	1.2034	-7471A	PPB		08 Mar 2017 12:08:03	U	SMPL
I2126-02 - 1	E2-Z7-WSW	933	0.2199	-7471A	PPB		08 Mar 2017 12:10:19	U	SMPL
CCV54 - 1	CCV54	22070	5.0122	-7471A	PPB		08 Mar 2017 12:12:35	U	SMPL
CCB54 - 1	CCB54	119	0.0353	-7471A	PPB		08 Mar 2017 12:14:51	U	SMPL
I2126-03 - 1	E2-Z7-BASE	67192	15.2425	-7471A	PPB		08 Mar 2017 12:17:07	U	SMPL
I2126-04 - 1	E2-AAZ-6-7-WSW	1129	0.2643	-7471A	PPB		08 Mar 2017 12:19:23	U	SMPL
I2126-05 - 1	E2-AA7-BASE	10716	2.4379	-7471A	PPB		08 Mar 2017 12:21:40	U	SMPL
I2133-02 - 1	HY-SB422A	24222	5.5001	-7471A	PPB		08 Mar 2017 12:23:57	U	SMPL
I2133-03 - 1	HY-SB422B	446	0.1094	-7471A	PPB		08 Mar 2017 12:26:14	U	SMPL
I2136-01 - 1	POSTEX-31-20170306	30771	6.9849	-7471A	PPB		08 Mar 2017 12:28:31	U	SMPL
I2137-04 - 1	38D-9-11	414	0.1022	-7471A	PPB		08 Mar 2017 12:30:48	U	SMPL
I2137-04DUP - 1	38D-9-11DUP	461	0.1128	-7471A	PPB		08 Mar 2017 12:36:09	U	SMPL
I2137-04MS - 1	38D-9-11MS	14817	3.3677	-7471A	PPB		08 Mar 2017 12:38:26	U	SMPL
I2137-04MSD - 1	38D-9-11MSD	16051	3.6475	-7471A	PPB		08 Mar 2017 12:40:42	U	SMPL
CCV55 - 1	CCV55	22569	5.1253	-7471A	PPB		08 Mar 2017 12:42:58	U	SMPL
CCB55 - 1	CCB55	118	0.0351	-7471A	PPB		08 Mar 2017 12:45:15	U	SMPL
I2137-04LX5 - 1		236	0.0618	-7471A	PPB		08 Mar 2017 12:47:31	U	SMPL
I2137-04A - 1		16532	3.7566	-7471A	PPB		08 Mar 2017 12:49:47	U	SMPL
I2126-03DLX5 - 1		13865	3.1519	-7471A	PPB		08 Mar 2017 12:52:04	U	SMPL
CCV56 - 1	CCV56	22106	5.0203	-7471A	PPB		08 Mar 2017 12:54:20	U	SMPL
CCB56 - 1	CCB56	114	0.0342	-7471A	PPB		08 Mar 2017 12:56:37	U	SMPL