



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86221

Temp.: 23.6 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	03/08/2017	09:15
2	CAL2	1	Water			7.02	03/08/2017	09:16
3	CAL3	1	Water			10.04	03/08/2017	09:17
4	ICV	1	Water			7.01	03/08/2017	09:19
5	CCV1	1	Water			2.02	03/08/2017	09:20
6	I2160-02	1	Solid	20.02	20	7.95	03/08/2017	09:21
7	I2160-02DUP	1	Solid	20.04	20	7.96	03/08/2017	09:22
8	I2160-04	1	Solid	20.02	20	7.76	03/08/2017	09:24
9	I2160-06	1	Solid	20.06	20	7.41	03/08/2017	09:25
10	I2160-08	1	Solid	20.01	20	8.04	03/08/2017	09:27
11	I2160-10	1	Solid	20.05	20	8.11	03/08/2017	09:29
12	CCV2	1	Water			12.02	03/08/2017	09:30

1386221

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2160

WorkList ID : 96283

Date : 3/8/2017 8:59:40 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2160-02	Corrosivity	Cool 4 deg C	ENTA04	K43	WC-B29-B35-001	03/07/2017	9045C
	Solid	I2160-04	Corrosivity	Cool 4 deg C	ENTA04	K43	WC-B44-B45-001	03/07/2017	9045C
	Solid	I2160-06	Corrosivity	Cool 4 deg C	ENTA04	K43	WC-B37-B40-001	03/07/2017	9045C
	Solid	I2160-08	Corrosivity	Cool 4 deg C	ENTA04	K43	WC-B41-B52-001	03/07/2017	9045C
	Solid	I2160-10	Corrosivity	Cool 4 deg C	ENTA04	K43	WC-B54-B56-001	03/07/2017	9045C

Date/Time 03/08/17 8:50 AM
 Received by: JA
 Relinquished by: JP

Date/Time 03/08/17 11:00 AM
 Received by: JP
 Relinquished by: JP